

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031120\
 Data File : PL057247.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2020 01:10
 Operator : AJ\MA
 Sample : L1826-02
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MC5B40

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 06:49:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030320.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 11 17:44:01 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.460	3.958	257.9E6	60941470	30.261	28.564
28)	SA Decachlor...	8.207	8.982	197.9E6	46903115	20.474m	22.075
Target Compounds							
16)	A 4,4'-DDD	6.142	6.768	14244486	9834950	1.471	4.037 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031120\
Data File : PL057247.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2020 01:10
Operator : AJ\MA
Sample : L1826-02
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
MC5B40

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 06:49:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030320.M
Quant Title : GC Extractables
QLast Update : Wed Mar 11 17:44:01 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.460 min
Delta R.T.: 0.000 min
Response: 257889909
Conc: 30.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
MC5B40




#1 Tetrachloro-m-xylene

R.T.: 3.958 min
Delta R.T.: 0.000 min
Response: 60941470
Conc: 28.56 ng/ml




#16 4,4'-DDD

R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 14244486
Conc: 1.47 ng/ml



#16 4,4'-DDD


R.T.: 6.768 min
Delta R.T.: -0.002 min
Response: 9834950
Conc: 4.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.207 min
Delta R.T.: -0.009 min
Response: 197886040
Conc: 20.47 ng/ml m

Instrument :
ECD_L
ClientSampleId :
MC5B40



#28 Decachlorobiphenyl

R.T.: 8.982 min
Delta R.T.: -0.002 min
Response: 46903115
Conc: 22.07 ng/ml