

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031219\
 Data File : PL045748.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Mar 2019 16:39
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 3/13/2019 9:36:53 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 01:20:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.362	3.981	358.4E6	111.2E6	52.103	51.925
28)	SA Decachlor...	8.044	9.038	385.7E6	112.7E6	46.098	49.309
Target Compounds							
2)	A alpha-BHC	3.792	4.371	592.6E6	163.6E6	53.351	53.829
3)	MA gamma-BHC...	4.069	4.655	545.3E6	150.9E6	53.955	53.420
4)	MA Heptachlor	4.357	5.162	559.4E6	155.8E6	53.012	52.370
5)	MB Aldrin	4.598	5.466	527.8E6	143.6E6	54.266	53.170
6)	B beta-BHC	4.322	4.827	224.6E6	68172154	52.392	53.681
7)	B delta-BHC	4.518	5.043	530.7E6	120.0E6	55.524	55.794
8)	B Heptachlor...	5.035	5.852	493.8E6	132.6E6	52.909	55.226
9)	A Endosulfan I	5.367	6.209	470.1E6	128.1E6	52.598	51.428
10)	B gamma-Chl...	5.258	6.087	517.8E6	140.0E6	52.910	52.169
11)	B alpha-Chl...	5.314	6.159	508.1E6	138.9E6	52.395	51.480
12)	B 4,4'-DDE	5.482	6.317	470.7E6	122.7E6	53.622	52.517
13)	MA Dieldrin	5.607	6.466	513.0E6	132.9E6	53.669	51.720
14)	MA Endrin	5.860	6.683	479.6E6	117.9E6	53.256	51.798
15)	B Endosulfa...	6.135	6.892	447.7E6	119.3E6	51.650	50.743m
16)	A 4,4'-DDD	5.993	6.807	405.5E6	103.2E6	55.389	53.673
17)	MA 4,4'-DDT	6.229	7.105	370.4E6	97169165	49.389	49.238
18)	B Endrin al...	6.302	7.017	367.7E6	101.4E6	50.304	49.620
19)	B Endosulfa...	6.513	7.239	407.3E6	112.1E6	50.111	49.843
20)	A Methoxychlor	6.774	7.564	204.3E6	57705089	46.511	47.013
21)	B Endrin ke...	6.996	7.710	468.1E6	118.4E6	50.808m	49.800
22)	Mirex	7.174	8.162	344.5E6	95604889	46.302	48.002

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031219\
Data File : PL045748.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Mar 2019 16:39
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
3/13/2019 9:36:53 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 01:20:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
Quant Title : GC Extractables
QLast Update : Wed Feb 27 17:00:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

