

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031325\
 Data File : PL094675.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Mar 2025 14:57
 Operator : AR\AJ
 Sample : Q1547-06
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OR-620-JB-COMP-02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 15:24:58 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 11 17:42:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.536	2.770	66225081	81463512	23.396	22.824
28)	SA Decachlor...	9.051	7.904	52969291	89312132	25.134	22.110
Target Compounds							
4)	MA Heptachlor	0.000	3.938	0	19063070	N.D.	3.618 #
22)	Mirex	0.000	7.009	0	5986073	N.D.	1.577 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031325\
Data File : PL094675.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2025 14:57
Operator : AR\AJ
Sample : Q1547-06
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
OR-620-JB-COMP-02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 15:24:58 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.536 min
Delta R.T.: -0.002 min
Response: 66225081
Conc: 23.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-620-JB-COMP-02




#1 Tetrachloro-m-xylene

R.T.: 2.770 min
Delta R.T.: -0.002 min
Response: 81463512
Conc: 22.82 ng/ml




#4 Heptachlor

R.T.: 0.000 min
Exp R.T. : 4.915 min
Response: 0
Conc: N.D.



#4 Heptachlor


R.T.: 3.938 min
Delta R.T.: -0.004 min
Response: 19063070
Conc: 3.62 ng/ml



#22 Mirex


R.T.: 0.000 min
Exp R.T.: 8.117 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
OR-620-JB-COMP-02




#22 Mirex

R.T.: 7.009 min
Delta R.T.: -0.006 min
Response: 5986073
Conc: 1.58 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.051 min
Delta R.T.: -0.005 min
Response: 52969291
Conc: 25.13 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.904 min
Delta R.T.: -0.003 min
Response: 89312132
Conc: 22.11 ng/ml