

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031419\
 Data File : PL045917.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Mar 2019 16:38
 Operator : AJ\SJ
 Sample : PB117820BS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB117820BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 01:09:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.361	3.980	110.4E6	36648533	16.043	17.113
28)	SA Decachlor...	8.043	9.037	143.1E6	40958874	17.106	17.926
Target Compounds							
2)	A alpha-BHC	3.791	4.370	510.6E6	146.4E6	45.966	48.165
3)	MA gamma-BHC...	4.069	4.655	470.7E6	135.8E6	46.567	48.065
4)	MA Heptachlor	4.357	5.162	480.9E6	140.8E6	45.574	47.329
5)	MB Aldrin	4.597	5.466	459.4E6	130.7E6	47.233	48.379
6)	B beta-BHC	4.322	4.827	196.3E6	62363866	45.808	49.107
7)	B delta-BHC	4.518	5.043	459.9E6	118.4E6	48.115	55.052
8)	B Heptachlor...	5.034	5.852	437.4E6	121.9E6	46.865	50.744
9)	A Endosulfan I	5.367	6.208	415.9E6	118.7E6	46.527	47.655
10)	B gamma-Chl...	5.257	6.086	450.4E6	128.7E6	46.015	47.967
11)	B alpha-Chl...	5.314	6.158	443.9E6	128.4E6	45.782	47.607
12)	B 4,4'-DDE	5.481	6.316	410.9E6	113.4E6	46.814	48.546
13)	MA Dieldrin	5.606	6.465	448.6E6	123.0E6	46.938	47.872
14)	MA Endrin	5.860	6.683	379.1E6	98400537	42.098	43.248
15)	B Endosulfa...	6.134	6.893	401.9E6	111.6E6	46.365	47.450
16)	A 4,4'-DDD	5.993	6.807	363.3E6	99014919	49.615	51.488
17)	MA 4,4'-DDT	6.228	7.105	302.4E6	84420242	40.318	42.778
18)	B Endrin al...	6.301	7.017	340.9E6	97566203	46.633	47.728
19)	B Endosulfa...	6.513	7.239	375.7E6	106.4E6	46.225	47.305
20)	A Methoxychlor	6.773	7.564	171.6E6	51646597	39.071	42.077
21)	B Endrin ke...	6.996	7.709	448.4E6	118.4E6	48.669	49.781
22)	Mirex	7.173	8.162	318.6E6	91166764	42.825	45.774

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031419\
Data File : PL045917.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Mar 2019 16:38
Operator : AJ\SJ
Sample : PB117820BS
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PB117820BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 01:09:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
Quant Title : GC Extractables
QLast Update : Wed Feb 27 17:00:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

