

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031419\
 Data File : PL045965.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2019 04:17
 Operator : AJ\SJ
 Sample : MIREX WATER IDOC 02
 Misc : PB117936
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MIREX WATER IDOC 02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 07:03:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.362	3.982	145.5E6	45164065	21.153	21.090
28)	SA Decachlor...	8.044	9.040	189.6E6	52330586	22.662	22.903
Target Compounds							
2)	A alpha-BHC	3.792	4.372	607.9E6	169.3E6	54.727	55.722
3)	MA gamma-BHC...	4.070	4.656	562.5E6	155.7E6	55.656	55.132
4)	MA Heptachlor	4.357	5.163	579.8E6	160.3E6	54.949	53.879
5)	MB Aldrin	4.598	5.468	553.1E6	150.3E6	56.869	55.659
6)	B beta-BHC	4.323	4.828	231.4E6	66872620	53.992	52.657
7)	B delta-BHC	4.519	5.044	552.0E6	145.2E6	57.751	67.516
8)	B Heptachlor...	5.036	5.854	529.4E6	143.4E6	56.723	59.700
9)	A Endosulfan I	5.368	6.210	501.9E6	133.6E6	56.151	53.653
10)	B gamma-Chl...	5.259	6.088	541.0E6	144.7E6	55.281	53.918
11)	B alpha-Chl...	5.315	6.160	530.9E6	142.8E6	54.750	52.939
12)	B 4,4'-DDE	5.482	6.318	497.2E6	131.5E6	56.635	56.300
13)	MA Dieldrin	5.607	6.467	543.8E6	141.3E6	56.900	54.999
14)	MA Endrin	5.861	6.684	463.5E6	112.5E6	51.470	49.437
15)	B Endosulfa...	6.135	6.895	485.1E6	126.6E6	55.956	53.826
16)	A 4,4'-DDD	5.994	6.809	436.6E6	114.2E6	59.636	59.387
17)	MA 4,4'-DDT	6.229	7.107	384.8E6	99461651	51.304	50.400
18)	B Endrin al...	6.302	7.019	411.4E6	108.5E6	56.280	53.071
19)	B Endosulfa...	6.514	7.241	456.1E6	119.3E6	56.120	53.052
20)	A Methoxychlor	6.774	7.565	210.1E6	55663393	47.845	45.350
21)	B Endrin ke...	6.997	7.711	546.1E6	135.3E6	59.273	56.867
22)	Mirex	7.174	8.164	379.4E6	100.4E6	50.991	50.389

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031419\
Data File : PL045965.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Mar 2019 04:17
Operator : AJ\SJ
Sample : MIREX WATER IDOC 02
Misc : PB117936
ALS Vial : 61 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
MIREX WATER IDOC 02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 07:03:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
Quant Title : GC Extractables
QLast Update : Wed Feb 27 17:00:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

