

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031419\
 Data File : PL045966.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Mar 2019 04:32
 Operator : AJ\SJ
 Sample : MIREX WATER IDOC 03
 Misc : PB117936
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MIREX WATER IDOC 03

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 15 07:03:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.362	3.982	145.8E6	45221438	21.193	21.116
28)	SA Decachlor...	8.044	9.041	189.6E6	52103768	22.663	22.803
Target Compounds							
2)	A alpha-BHC	3.792	4.372	609.0E6	168.4E6	54.826	55.426
3)	MA gamma-BHC...	4.070	4.657	560.7E6	154.8E6	55.478	54.820
4)	MA Heptachlor	4.357	5.163	578.2E6	159.3E6	54.794	53.538
5)	MB Aldrin	4.598	5.468	552.7E6	149.9E6	56.825	55.501
6)	B beta-BHC	4.322	4.829	230.9E6	66582511	53.863	52.429
7)	B delta-BHC	4.518	5.044	550.3E6	144.3E6	57.573	67.090
8)	B Heptachlor...	5.035	5.854	525.7E6	141.8E6	56.329	59.050
9)	A Endosulfan I	5.367	6.211	500.3E6	132.9E6	55.976	53.377
10)	B gamma-Chl...	5.258	6.088	538.6E6	144.1E6	55.035	53.728
11)	B alpha-Chl...	5.315	6.161	528.9E6	142.1E6	54.545	52.680
12)	B 4,4'-DDE	5.482	6.318	495.3E6	131.4E6	56.424	56.260
13)	MA Dieldrin	5.607	6.467	542.4E6	140.5E6	56.753	54.683
14)	MA Endrin	5.861	6.685	457.5E6	110.6E6	50.802	48.615
15)	B Endosulfa...	6.135	6.895	483.6E6	125.5E6	55.787	53.347
16)	A 4,4'-DDD	5.994	6.809	437.0E6	114.3E6	59.691	59.454
17)	MA 4,4'-DDT	6.229	7.107	382.0E6	97735362	50.935	49.525
18)	B Endrin al...	6.302	7.018	410.6E6	108.0E6	56.166	52.812
19)	B Endosulfa...	6.514	7.241	453.2E6	118.5E6	55.762	52.680
20)	A Methoxychlor	6.774	7.565	208.3E6	55072623	47.436	44.869
21)	B Endrin ke...	6.998	7.711	546.7E6	134.9E6	59.347	56.720
22)	Mirex	7.174	8.164	377.7E6	99741767	50.771	50.079

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031419\
Data File : PL045966.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Mar 2019 04:32
Operator : AJ\SJ
Sample : MIREX WATER IDOC 03
Misc : PB117936
ALS Vial : 62 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
MIREX WATER IDOC 03

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 15 07:03:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
Quant Title : GC Extractables
QLast Update : Wed Feb 27 17:00:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

