

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031420\
 Data File : PL057336.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Mar 2020 17:09
 Operator : AJ\MA
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 14 00:02:07 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031420.M
 Quant Title : GC Extractables
 QLast Update : Sat Mar 14 00:01:37 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.458	3.957	468.9E6	108.4E6	50.000	50.000
28)	SA Decachlor...	8.207	8.980	514.1E6	107.9E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.895	4.345	821.4E6	163.0E6	50.000	50.000
3)	MA gamma-BHC...	4.179	4.627	744.1E6	158.6E6	50.000	50.000
4)	MA Heptachlor	4.479	5.129	752.8E6	156.4E6	50.000	50.000
5)	MB Aldrin	4.727	5.431	706.4E6	152.7E6	50.000	50.000
6)	B beta-BHC	4.431	4.800	290.8E6	71311340	50.000	50.000
7)	B delta-BHC	4.635	5.013	725.7E6	122.7E6	50.000	50.000
8)	B Heptachlo...	5.170	5.815	648.8E6	136.2E6	50.000	50.000
9)	A Endosulfan I	5.509	6.170	602.5E6	140.9E6	50.000	50.000
10)	B gamma-Chl...	5.397	6.049	659.9E6	152.9E6	50.000	50.000
11)	B alpha-Chl...	5.455	6.120	652.2E6	154.3E6	50.000	50.000
12)	B 4,4'-DDE	5.619	6.280	634.2E6	142.7E6	50.000	50.000
13)	MA Dieldrin	5.751	6.425	645.1E6	147.3E6	50.000	50.000
14)	MA Endrin	6.009	6.642	582.5E6	120.2E6	50.000	50.000
15)	B Endosulfa...	6.282	6.852	550.8E6	127.3E6	50.000	50.000
16)	A 4,4'-DDD	6.134	6.768	501.9E6	114.7E6	50.000	50.000
17)	MA 4,4'-DDT	6.375	7.065	520.0E6	115.2E6	50.000	50.000
18)	B Endrin al...	6.450	6.976	463.3E6	109.8E6	50.000	50.000
19)	B Endosulfa...	6.663	7.199	515.5E6	120.4E6	50.000	50.000
20)	A Methoxychlor	6.920	7.524	271.4E6	64735875	50.000	50.000
21)	B Endrin ke...	7.154	7.666	575.5E6	128.3E6	50.000	50.000
22)	Mirex	7.341	8.116	400.0E6	93092238	50.000	50.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031420\
Data File : PL057336.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Mar 2020 17:09
Operator : AJ\MA
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 14 00:02:07 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031420.M
Quant Title : GC Extractables
QLast Update : Sat Mar 14 00:01:37 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

