

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031519\
 Data File : PL046052.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Mar 2019 08:34
 Operator : AJ\SJ
 Sample : K1930-05
 Misc :
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-4

Manual Integrations
 APPROVED

Sohil
 3/18/2019 11:40:51 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 18 01:21:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 27 17:00:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.361	3.981	94969016	29887870	13.805	13.956
28)	SA Decachlor...	8.043	9.039	87183440	25514013	10.419	11.166
Target Compounds							
12)	B 4,4'-DDE	5.481	6.318	718.7E6	192.9E6	81.874	82.581
13)	MA Dieldrin	5.606	6.465	442.2E6	115.0E6	46.266	44.753m
16)	A 4,4'-DDD	5.993	6.806	137.5E6	40966271	18.776	21.302m
17)	MA 4,4'-DDT	6.228	7.105	382.9E6	100.4E6	51.049	50.867

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031519\
Data File : PL046052.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Mar 2019 08:34
Operator : AJ\SJ
Sample : K1930-05
Misc :
ALS Vial : 73 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-4

Manual Integrations
APPROVED

Sohil
3/18/2019 11:40:51 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 18 01:21:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL022719.M
Quant Title : GC Extractables
QLast Update : Wed Feb 27 17:00:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

