

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX031621.M
 Title : GC Extractables
 Last Update : Wed Mar 17 04:52:08 2021
 Response Via : Initial Calibration

Calibration Files

500 =PL065341.D 1000 =PL065339.D 250 =PL065342.D
 750 =PL065340.D 100 =PL065343.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.390	2.403	2.366	2.393	2.375	2.385	E6	0.63
2)	Toxaphene-1	1.554	1.504	1.659	1.497	1.660	1.575	E4	5.10
3)	Toxaphene-2	1.484	1.594	1.559	1.545	1.452	1.527	E4	3.76
4)	Toxaphene-3	4.046	4.187	4.030	4.197	3.639	4.020	E4	5.64
5)	Toxaphene-4	2.783	2.950	2.706	2.973	2.383	2.759	E4	8.64
6)	Toxaphene-5	4.844	4.929	4.687	4.915	4.424	4.760	E4	4.43
7) SA	Decachlorobiphenyl	2.366	2.261	2.459	2.317	2.557	2.392	E6	4.91

Signal #2 Calibration Files

500 =PL065341.D 1000 =PL065339.D 250 =PL065342.D
 750 =PL065340.D 100 =PL065343.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.769	2.733	2.841	2.738	2.860	2.788	E6	2.11
2)	Toxaphene-1	1.692	1.596	1.662	1.645	1.611	1.641	E4	2.35
3)	Toxaphene-2	2.028	1.907	2.047	1.962	1.858	1.961	E4	4.07
4)	Toxaphene-3	4.392	4.197	4.622	4.224	4.520	4.391	E4	4.20
5)	Toxaphene-4	3.777	3.650	3.800	3.754	3.514	3.699	E4	3.19
6)	Toxaphene-5	3.239	3.077	3.245	3.108	2.844	3.102	E4	5.26
7) SA	Decachlorobiphenyl	1.975	1.881	2.032	1.930	2.040	1.972	E6	3.44

(#) = Out of Range