

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031621\
 Data File : PL065344.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Mar 2021 19:57
 Operator : AR\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL031621

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 17 05:40:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031621.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 17 04:35:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.293	3.883	122.6E6	141.9E6	50.406	49.159
28)	SA Decachlor...	7.937	8.890	121.9E6	100.9E6	49.638	49.605
Target Compounds							
2)	A alpha-BHC	3.716	4.274	197.7E6	215.4E6	51.263	49.658
3)	MA gamma-BHC...	3.990	4.557	178.0E6	193.7E6	50.715	48.947
4)	MA Heptachlor	4.276	5.051	172.4E6	182.2E6	50.342	48.678
5)	MB Aldrin	4.514	5.352	167.3E6	177.3E6	51.024	49.465
6)	B beta-BHC	4.240	4.739	73796632	87230195	50.319	51.434
7)	B delta-BHC	4.434	4.952	176.0E6	182.0E6	51.249	50.215
8)	B Heptachlo...	4.947	5.741	156.9E6	157.2E6	50.305	49.332
9)	A Endosulfan I	5.275	6.096	144.2E6	146.6E6	50.440	49.110
10)	B gamma-Chl...	5.169	5.974	156.3E6	156.8E6	50.499	49.210
11)	B alpha-Chl...	5.224	6.044	153.2E6	153.0E6	50.437	49.177
12)	B 4,4'-DDE	5.391	6.206	136.7E6	147.9E6	51.065	49.622
13)	MA Dieldrin	5.513	6.353	148.1E6	154.8E6	50.839	49.453
14)	MA Endrin	5.763	6.571	127.0E6	125.7E6	48.258	48.007
15)	B Endosulfa...	6.035	6.787	127.7E6	123.0E6	50.294	49.492
16)	A 4,4'-DDD	5.898	6.700	119.2E6	124.9E6	53.566	49.482
17)	MA 4,4'-DDT	6.132	6.995	96172423	109.2E6	49.941	49.024
18)	B Endrin al...	6.202	6.912	100.5E6	98017807	50.036	49.755
19)	B Endosulfa...	6.412	7.134	132.6E6	122.3E6	50.021	48.698
20)	A Methoxychlor	6.677	7.459	55276896	54196289	49.512	48.241
21)	B Endrin ke...	6.891	7.604	147.2E6	135.8E6	51.751	49.879
22)	Mirex	7.064	8.043	109.2E6	88849562	49.995	49.181

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031621\
Data File : PL065344.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Mar 2021 19:57
Operator : AR\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL031621

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 17 05:40:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031621.M
Quant Title : GC Extractables
QLast Update : Wed Mar 17 04:35:29 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

