

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031623\
 Data File : PL081522.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Mar 2023 21:07
 Operator : AR\AJ
 Sample : 01932-19
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 JPP-6C-031423

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/17/2023
 Supervised By : Ankita Jodhani 03/17/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 22:07:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 03 01:48:45 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.329	2.661	19269428	19106028	8.919m	9.441
2) SA Decachlor...	8.754	7.781	15310974	19231650	10.043	10.040

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031623\
Data File : PL081522.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Mar 2023 21:07
Operator : AR\AJ
Sample : 01932-19
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
JPP-6C-031423


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/17/2023
Supervised By : Ankita Jodhani 03/17/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 22:07:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL030223.M
Quant Title : GC Extractables
QLast Update : Fri Mar 03 01:48:45 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.329 min
Delta R.T.: -0.006 min
Response: 19269428
Conc: 8.92 ng/ml

Instrument :
ECD_L
ClientSampleId :
JPP-6C-031423


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/17/2023
Supervised By :Ankita Jodhani 03/17/2023




#1 Tetrachloro-m-xylene

R.T.: 2.661 min
Delta R.T.: -0.006 min
Response: 19106028
Conc: 9.44 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.754 min
Delta R.T.: -0.011 min
Response: 15310974
Conc: 10.04 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.781 min
Delta R.T.: -0.012 min
Response: 19231650
Conc: 10.04 ng/ml