

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031822\
 Data File : PL073489.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Mar 2022 20:21
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 01:17:49 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031122.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 14 04:12:46 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.642	5.492	14604519	82734747	45.598	50.410
28)	SA Decachlor...	10.416	11.332	22489368	80423197	44.227	44.975
Target Compounds							
2)	A alpha-BHC	5.350	6.055	19030279	117.9E6	46.167	52.406
3)	MA gamma-BHC...	5.774	6.444	19034767	110.6E6	45.632	51.781
4)	MA Heptachlor	6.186	7.103	20488904	120.7E6	45.410	51.513
5)	MB Aldrin	6.511	7.471	18617965	109.0E6	45.061	52.033
6)	B beta-BHC	6.154	6.680	9618731	51010877	46.716	48.937
7)	B delta-BHC	6.416	6.947	18544911	109.4E6	46.172	50.779
8)	B Heptachlo...	7.088	7.927	19231317	102.7E6	44.622	51.421
9)	A Endosulfan I	7.491	8.327	18317736	92477971	44.654	50.662
10)	B gamma-Chl...	7.365	8.195	19408714	106.4E6	44.383	52.113
11)	B alpha-Chl...	7.435	8.277	19620740	102.0E6	44.398	51.007
12)	B 4,4'-DDE	7.649	8.461	15962626	91018985	44.484	50.781
13)	MA Dieldrin	7.777	8.612	19134720	92251748	45.928	50.008
14)	MA Endrin	8.069	8.846	17518433	75807699	45.150	49.244
15)	B Endosulfa...	8.379	9.070	17257052	75434490	44.150	49.116
16)	A 4,4'-DDD	8.234	8.989	14274962	74694691	43.958	51.171
17)	MA 4,4'-DDT	8.496	9.307	16358573	77728828	43.256	47.554
18)	B Endrin al...	8.568	9.203	13972676	53099139	44.183	43.300
19)	B Endosulfa...	8.799	9.440	18605015	75669788	44.044	48.012
20)	A Methoxychlor	9.092	9.786	10894376	41172598	42.467	45.693
21)	B Endrin ke...	9.318	9.928	22090610	79090371	45.154	47.916
22)	Mirex	9.506	10.404	20309943	64108012	43.069	47.192
23)	Chlordane-1	0.000	6.854	0	1300592	N.D.	19.039 #
24)	Chlordane-2	0.000	7.471f	0	109.0E6	N.D.	1516.647 #
25)	Chlordane-3	7.365	8.195	19408714	106.4E6	475.306	392.560
26)	Chlordane-4	7.435	8.277	19620740	102.0E6	442.698	307.966 #
27)	Chlordane-5	8.379	0.000	17257052	0	1053.847	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031822\
Data File : PL073489.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Mar 2022 20:21
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 01:17:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031122.M
Quant Title : GC Extractables
QLast Update : Mon Mar 14 04:12:46 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.642 min
Delta R.T.: -0.001 min
Response: 14604519
Conc: 45.60 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 5.492 min
Delta R.T.: 0.000 min
Response: 82734747
Conc: 50.41 ng/ml




#2 alpha-BHC

R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 19030279
Conc: 46.17 ng/ml



#2 alpha-BHC


R.T.: 6.055 min
Delta R.T.: 0.001 min
Response: 117894871
Conc: 52.41 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 5.774 min
Delta R.T.: -0.001 min
Response: 19034767
Conc: 45.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 6.444 min
Delta R.T.: 0.001 min
Response: 110556957
Conc: 51.78 ng/ml




#4 Heptachlor

R.T.: 6.186 min
Delta R.T.: -0.001 min
Response: 20488904
Conc: 45.41 ng/ml



#4 Heptachlor


R.T.: 7.103 min
Delta R.T.: 0.001 min
Response: 120699142
Conc: 51.51 ng/ml



#5 Aldrin


R.T.: 6.511 min
Delta R.T.: -0.001 min
Response: 18617965
Conc: 45.06 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 7.471 min
Delta R.T.: 0.001 min
Response: 109007025
Conc: 52.03 ng/ml




#6 beta-BHC

R.T.: 6.154 min
Delta R.T.: 0.000 min
Response: 9618731
Conc: 46.72 ng/ml



#6 beta-BHC


R.T.: 6.680 min
Delta R.T.: 0.000 min
Response: 51010877
Conc: 48.94 ng/ml



#7 delta-BHC


R.T.: 6.416 min
Delta R.T.: 0.000 min
Response: 18544911
Conc: 46.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 6.947 min
Delta R.T.: 0.001 min
Response: 109354980
Conc: 50.78 ng/ml




#8 Heptachlor epoxide

R.T.: 7.088 min
Delta R.T.: -0.002 min
Response: 19231317
Conc: 44.62 ng/ml



#8 Heptachlor epoxide


R.T.: 7.927 min
Delta R.T.: 0.000 min
Response: 102704611
Conc: 51.42 ng/ml



#9 Endosulfan I


R.T.: 7.491 min
Delta R.T.: -0.002 min
Response: 18317736
Conc: 44.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 8.327 min
Delta R.T.: 0.000 min
Response: 92477971
Conc: 50.66 ng/ml




#10 gamma-Chlordane

R.T.: 7.365 min
Delta R.T.: -0.002 min
Response: 19408714
Conc: 44.38 ng/ml



#10 gamma-Chlordane


R.T.: 8.195 min
Delta R.T.: 0.001 min
Response: 106435740
Conc: 52.11 ng/ml



#11 alpha-Chlordane


R.T.: 7.435 min
Delta R.T.: -0.002 min
Response: 19620740
Conc: 44.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 8.277 min
Delta R.T.: 0.000 min
Response: 102014711
Conc: 51.01 ng/ml




#12 4,4'-DDE

R.T.: 7.649 min
Delta R.T.: -0.002 min
Response: 15962626
Conc: 44.48 ng/ml



#12 4,4'-DDE


R.T.: 8.461 min
Delta R.T.: 0.000 min
Response: 91018985
Conc: 50.78 ng/ml



#13 Dieldrin


R.T.: 7.777 min
Delta R.T.: -0.002 min
Response: 19134720
Conc: 45.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 8.612 min
Delta R.T.: 0.000 min
Response: 92251748
Conc: 50.01 ng/ml




#14 Endrin

R.T.: 8.069 min
Delta R.T.: -0.002 min
Response: 17518433
Conc: 45.15 ng/ml



#14 Endrin


R.T.: 8.846 min
Delta R.T.: 0.000 min
Response: 75807699
Conc: 49.24 ng/ml



#15 Endosulfan II


R.T.: 8.379 min
Delta R.T.: -0.003 min
Response: 17257052
Conc: 44.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 9.070 min
Delta R.T.: 0.001 min
Response: 75434490
Conc: 49.12 ng/ml




#16 4,4'-DDD

R.T.: 8.234 min
Delta R.T.: -0.003 min
Response: 14274962
Conc: 43.96 ng/ml



#16 4,4'-DDD


R.T.: 8.989 min
Delta R.T.: 0.000 min
Response: 74694691
Conc: 51.17 ng/ml



#17 4,4'-DDT


R.T.: 8.496 min
Delta R.T.: -0.003 min
Response: 16358573
Conc: 43.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 9.307 min
Delta R.T.: 0.000 min
Response: 77728828
Conc: 47.55 ng/ml




#18 Endrin aldehyde

R.T.: 8.568 min
Delta R.T.: -0.002 min
Response: 13972676
Conc: 44.18 ng/ml



#18 Endrin aldehyde


R.T.: 9.203 min
Delta R.T.: 0.000 min
Response: 53099139
Conc: 43.30 ng/ml



#19 Endosulfan Sulfate


R.T.: 8.799 min
Delta R.T.: -0.002 min
Response: 18605015
Conc: 44.04 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 9.440 min
Delta R.T.: 0.000 min
Response: 75669788
Conc: 48.01 ng/ml




#20 Methoxychlor

R.T.: 9.092 min
Delta R.T.: -0.003 min
Response: 10894376
Conc: 42.47 ng/ml



#20 Methoxychlor


R.T.: 9.786 min
Delta R.T.: 0.000 min
Response: 41172598
Conc: 45.69 ng/ml



#21 Endrin ketone


R.T.: 9.318 min
Delta R.T.: -0.002 min
Response: 22090610
Conc: 45.15 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 9.928 min
Delta R.T.: 0.000 min
Response: 79090371
Conc: 47.92 ng/ml




#22 Mirex

R.T.: 9.506 min
Delta R.T.: -0.003 min
Response: 20309943
Conc: 43.07 ng/ml



#22 Mirex


R.T.: 10.404 min
Delta R.T.: 0.000 min
Response: 64108012
Conc: 47.19 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 5.976 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#23 Chlordane-1

R.T.: 6.854 min
Delta R.T.: -0.009 min
Response: 1300592
Conc: 19.04 ng/ml




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T.: 6.665 min
Response: 0
Conc: N.D.



#24 Chlordane-2


R.T.: 7.471 min
Delta R.T.: 0.025 min
Response: 109007025
Conc: 1516.65 ng/ml



#25 Chlordane-3


R.T.: 7.365 min
Delta R.T.: -0.003 min
Response: 19408714
Conc: 475.31 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#25 Chlordane-3

R.T.: 8.195 min
Delta R.T.: 0.000 min
Response: 106435740
Conc: 392.56 ng/ml




#26 Chlordane-4

R.T.: 7.435 min
Delta R.T.: -0.002 min
Response: 19620740
Conc: 442.70 ng/ml



#26 Chlordane-4


R.T.: 8.277 min
Delta R.T.: -0.004 min
Response: 102014711
Conc: 307.97 ng/ml



#27 Chlordane-5


R.T.: 8.379 min
Delta R.T.: -0.007 min
Response: 17257052
Conc: 1053.85 ng/m

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 9.149 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 10.416 min
Delta R.T.: -0.005 min
Response: 22489368
Conc: 44.23 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.332 min
Delta R.T.: 0.000 min
Response: 80423197
Conc: 44.98 ng/ml