

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031921\
 Data File : PL065436.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Mar 2021 13:16
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 19 14:04:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 19 02:14:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.288	3.788	134.3E6	131.4E6	56.559	47.886
28)	SA Decachlor...	7.929	8.739	113.8E6	101.6E6	47.121	51.279
Target Compounds							
2)	A alpha-BHC	3.712	4.173	214.5E6	190.8E6	57.367	46.954
3)	MA gamma-BHC...	3.986	4.452	192.0E6	171.7E6	56.140	47.033
4)	MA Heptachlor	4.271	4.938	164.3E6	141.7E6	50.206	42.487
5)	MB Aldrin	4.509	5.235	178.4E6	148.8E6	56.311	45.422
6)	B beta-BHC	4.236	4.634	79176803	76146688	55.991	44.201
7)	B delta-BHC	4.429	4.842	186.3E6	162.6E6	56.004	54.405
8)	B Heptachlo...	4.941	5.621	162.0E6	147.0E6	54.786	50.863
9)	A Endosulfan I	5.269	5.971	145.9E6	122.9E6	53.514	46.667
10)	B gamma-Chl...	5.162	5.852	158.4E6	135.7E6	53.638	48.645
11)	B alpha-Chl...	5.218	5.922	155.8E6	131.4E6	53.532	47.345
12)	B 4,4'-DDE	5.385	6.085	141.5E6	114.6E6	53.788	46.024
13)	MA Dieldrin	5.506	6.227	148.7E6	123.6E6	53.107	45.452
14)	MA Endrin	5.755	6.443	126.6E6	100.8E6	50.707	45.637
15)	B Endosulfa...	6.028	6.658	120.9E6	100.7E6	49.370	44.627
16)	A 4,4'-DDD	5.891	6.576	118.4E6	99745958	53.591	45.349
17)	MA 4,4'-DDT	6.125	6.869	97319642	89514036	51.417	45.606
18)	B Endrin al...	6.195	6.783	89908189	78847261	46.821	45.007
19)	B Endosulfa...	6.405	7.004	126.0E6	104.7E6	50.985	47.707
20)	A Methoxychlor	6.670	7.334	54230425	48059568	47.300	46.018
21)	B Endrin ke...	6.883	7.470	134.6E6	115.5E6	50.081	48.935
22)	Mirex	7.055	7.906	99785082	84355311	49.003	50.680

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031921\
Data File : PL065436.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Mar 2021 13:16
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 19 14:04:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
Quant Title : GC Extractables
QLast Update : Fri Mar 19 02:14:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

