

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032125\
 Data File : PL094782.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Mar 2025 12:37
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 21 13:16:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 11 17:42:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.536	2.770	132.6E6	181.5E6	46.861	50.844
2)	SA Decachlor...	9.055	7.904	114.0E6	212.8E6	54.107	52.679
Target Compounds							
2)	A alpha-BHC	3.992	3.272	194.5E6	279.5E6	46.842	51.846
3)	MA gamma-BHC...	4.325	3.602	187.8E6	264.0E6	47.055	51.364
4)	MA Heptachlor	4.913	3.939	181.3E6	266.6E6	46.712	50.602
5)	MB Aldrin	5.254	4.219	170.9E6	249.4E6	46.280	51.150
6)	B beta-BHC	4.524	3.902	85349089	114.6E6	46.254	51.598
7)	B delta-BHC	4.771	4.131	188.6E6	263.1E6	48.429	52.601
8)	B Heptachlo...	5.681	4.722	158.5E6	234.8E6	47.389	51.287
9)	A Endosulfan I	6.068	5.091	148.7E6	225.0E6	48.426	51.275
10)	B gamma-Chl...	5.938	4.972	160.7E6	247.6E6	47.697	51.273
11)	B alpha-Chl...	6.017	5.035	158.3E6	243.5E6	48.010	51.013
12)	B 4,4'-DDE	6.191	5.224	157.1E6	249.6E6	53.397	53.688
13)	MA Dieldrin	6.343	5.355	155.6E6	252.8E6	48.650	52.104
14)	MA Endrin	6.572	5.631	136.2E6	230.2E6	49.129	52.752
15)	B Endosulfa...	6.793	5.926	133.1E6	223.7E6	49.038	51.682
16)	A 4,4'-DDD	6.709	5.779	116.6E6	200.7E6	53.830	55.803
17)	MA 4,4'-DDT	7.023	6.029	128.1E6	224.3E6	53.844	55.620
18)	B Endrin al...	6.923	6.105	102.0E6	168.6E6	48.339	50.109
19)	B Endosulfa...	7.158	6.329	122.3E6	217.9E6	50.270	53.497
20)	A Methoxychlor	7.499	6.604	67830701	120.1E6	56.664	56.616
21)	B Endrin ke...	7.643	6.833	136.3E6	252.3E6	51.579	52.861
22)	Mirex	8.115	7.013	101.5E6	192.0E6	49.115	50.592

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL031215\
Data File : PL094782.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Mar 2025 12:37
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 13:16:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

