

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032221\
 Data File : PL065462.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Mar 2021 09:42
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 3/23/2021 11:18:27 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 02:00:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 19 02:14:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.291	3.786	138.7E6	135.0E6	58.384	49.196
28)	SA Decachlor...	7.929	8.736	117.3E6	104.2E6	48.531	52.546
Target Compounds							
2)	A alpha-BHC	3.715	4.171	220.4E6	195.3E6	58.946	48.062
3)	MA gamma-BHC...	3.988	4.449	197.8E6	177.1E6	57.834	48.508m
4)	MA Heptachlor	4.273	4.936	177.0E6	154.9E6	54.076	46.443
5)	MB Aldrin	4.511	5.233	183.5E6	154.5E6	57.914	47.172
6)	B beta-BHC	4.238	4.633	82139179	78626430	58.085	45.641
7)	B delta-BHC	4.431	4.840	193.6E6	167.8E6	58.221	56.140
8)	B Heptachlo...	4.944	5.619	167.7E6	143.8E6	56.707	49.743
9)	A Endosulfan I	5.270	5.969	150.7E6	128.9E6	55.272	48.946
10)	B gamma-Chl...	5.164	5.850	163.7E6	139.5E6	55.457	49.979
11)	B alpha-Chl...	5.220	5.919	160.4E6	137.0E6	55.122	49.378
12)	B 4,4'-DDE	5.387	6.084	145.8E6	122.1E6	55.422	49.034
13)	MA Dieldrin	5.508	6.225	153.9E6	130.7E6	54.974	48.045
14)	MA Endrin	5.757	6.440	128.6E6	104.2E6	51.484	47.193
15)	B Endosulfa...	6.029	6.656	128.6E6	107.3E6	52.490	47.546
16)	A 4,4'-DDD	5.892	6.574	126.3E6	105.6E6	57.136	47.992
17)	MA 4,4'-DDT	6.126	6.867	102.5E6	92201261	54.131	46.975
18)	B Endrin al...	6.196	6.781	95196454	83445380	49.575	47.632
19)	B Endosulfa...	6.406	7.002	132.8E6	110.0E6	53.726	50.124
20)	A Methoxychlor	6.671	7.332	56630681	49275355	49.394	47.182
21)	B Endrin ke...	6.884	7.467	142.9E6	122.3E6	53.161	51.822
22)	Mirex	7.056	7.904	103.9E6	87347028	51.015	52.478

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032221\
Data File : PL065462.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Mar 2021 09:42
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
3/23/2021 11:18:27 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 02:00:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
Quant Title : GC Extractables
QLast Update : Fri Mar 19 02:14:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

