

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032221\
 Data File : PL065464.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Mar 2021 11:12
 Operator : AR\AJ
 Sample : M1721-03
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TR-08-031821

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 02:01:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 19 02:14:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.293	3.786	34941260	35644103	14.713	12.986
28)	SA Decachlor...	7.934	8.737	34887326	30164949	14.440	15.218
Target Compounds							
12)	B 4,4'-DDE	5.391	6.084	2702638	1927926	1.027	0.774



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032221\
Data File : PL065464.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Mar 2021 11:12
Operator : AR\AJ
Sample : M1721-03
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TR-08-031821

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 02:01:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
Quant Title : GC Extractables
QLast Update : Fri Mar 19 02:14:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.293 min
Delta R.T.: 0.004 min
Response: 34941260
Conc: 14.71 ng/ml

Instrument :
ECD_L
ClientSampleId :
TR-08-031821




#1 Tetrachloro-m-xylene

R.T.: 3.786 min
Delta R.T.: -0.004 min
Response: 35644103
Conc: 12.99 ng/ml




#12 4,4'-DDE

R.T.: 5.391 min
Delta R.T.: 0.003 min
Response: 2702638
Conc: 1.03 ng/ml



#12 4,4'-DDE


R.T.: 6.084 min
Delta R.T.: -0.005 min
Response: 1927926
Conc: 0.77 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.934 min
Delta R.T.: 0.000 min
Response: 34887326
Conc: 14.44 ng/ml

Instrument :
ECD_L
ClientSampleId :
TR-08-031821



#28 Decachlorobiphenyl

R.T.: 8.737 min
Delta R.T.: -0.006 min
Response: 30164949
Conc: 15.22 ng/ml