

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032221\
 Data File : PL065473.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Mar 2021 15:47
 Operator : AR\AJ
 Sample : M1718-03
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 BACKFILL-ASHCROFT-REAR-05

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 02:04:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 19 02:14:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.286	3.785	29974888	30105363	12.622	10.968
28)	SA Decachlor...	7.926	8.734	32703756	27446174	13.536	13.847

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032221\
Data File : PL065473.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Mar 2021 15:47
Operator : AR\AJ
Sample : M1718-03
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
BACKFILL-ASHCROFT-REAR-05

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 02:04:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
Quant Title : GC Extractables
QLast Update : Fri Mar 19 02:14:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.286 min
Delta R.T.: -0.003 min
Response: 29974888
Conc: 12.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
BACKFILL-ASHCROFT-REAR-05




#1 Tetrachloro-m-xylene

R.T.: 3.785 min
Delta R.T.: -0.005 min
Response: 30105363
Conc: 10.97 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.926 min
Delta R.T.: -0.007 min
Response: 32703756
Conc: 13.54 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.734 min
Delta R.T.: -0.008 min
Response: 27446174
Conc: 13.85 ng/ml