

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032223\
 Data File : PL081605.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Mar 2023 18:15
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 03/23/2023
 Supervised By :Ankita Jodhani 03/23/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 00:47:07 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 23 00:46:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.329	2.659	228.2E6	202.2E6	97.497	99.283
28)	SA Decachlor...	8.749	7.776	161.1E6	183.1E6	94.929	97.047
Target Compounds							
2)	A alpha-BHC	3.779	3.158	363.6E6	315.9E6	100.539	102.225
3)	MA gamma-BHC...	4.106	3.486	330.2E6	292.1E6	99.079	101.411
4)	MA Heptachlor	4.684	3.819	317.4E6	297.2E6	96.860m	99.728
5)	MB Aldrin	5.022	4.097	292.1E6	273.7E6	98.493	101.185
6)	B beta-BHC	4.308	3.785	138.4E6	115.9E6	97.544	99.352
7)	B delta-BHC	4.550	4.012	309.7E6	276.7E6	97.449	102.071
8)	B Heptachlo...	5.448	4.599	260.5E6	246.2E6	97.783	99.560
9)	A Endosulfan I	5.829	4.966	237.2E6	229.9E6	96.571	100.319
10)	B gamma-Chl...	5.703	4.848	261.0E6	249.5E6	97.349	100.374
11)	B alpha-Chl...	5.781	4.911	255.7E6	247.0E6	96.941	99.790
12)	B 4,4'-DDE	5.959	5.101	232.1E6	208.6E6	98.755	101.984
13)	MA Dieldrin	6.103	5.231	244.5E6	237.0E6	97.582	101.746
14)	MA Endrin	6.330	5.505	212.8E6	211.7E6	99.065	101.187
15)	B Endosulfa...	6.552	5.801	197.4E6	187.6E6	96.768	99.488
16)	A 4,4'-DDD	6.474	5.656	183.1E6	164.7E6	97.152	102.072
17)	MA 4,4'-DDT	6.785	5.905	200.8E6	182.7E6	97.232	101.621
18)	B Endrin al...	6.682	5.982	156.0E6	152.3E6	95.557	98.650
19)	B Endosulfa...	6.916	6.204	201.4E6	193.6E6	96.103	99.342
20)	A Methoxychlor	7.265	6.484	102.0E6	97487347	94.703	98.524
21)	B Endrin ke...	7.395	6.709	206.2E6	217.1E6	96.545	100.145
22)	Mirex	7.860	6.881	157.4E6	181.1E6	94.290m	96.498

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032223\
Data File : PL081605.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Mar 2023 18:15
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/23/2023
Supervised By :Ankita Jodhani 03/23/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 00:47:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
Quant Title : GC Extractables
QLast Update : Thu Mar 23 00:46:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

