

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032223\
 Data File : PL081606.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Mar 2023 18:29
 Operator : AR\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 00:50:15 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 23 00:50:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.329	2.659	172.2E6	152.5E6	74.048	74.904
28)	SA Decachlor...	8.749	7.776	125.2E6	140.8E6	74.182	74.756
Target Compounds							
2)	A alpha-BHC	3.779	3.158	270.6E6	234.1E6	74.884	75.509
3)	MA gamma-BHC...	4.106	3.486	244.6E6	216.6E6	73.918	75.141
4)	MA Heptachlor	4.686	3.819	233.5E6	221.5E6	72.461	74.563
5)	MB Aldrin	5.023	4.097	219.0E6	205.0E6	74.220	75.522
6)	B beta-BHC	4.309	3.785	105.4E6	88341172	74.534	75.484
7)	B delta-BHC	4.551	4.012	232.4E6	206.0E6	73.736	75.644
8)	B Heptachlo...	5.449	4.598	198.5E6	186.7E6	74.667	75.320
9)	A Endosulfan I	5.830	4.966	180.5E6	169.7E6	73.986	74.384
10)	B gamma-Chl...	5.703	4.847	196.6E6	186.5E6	73.863	75.002
11)	B alpha-Chl...	5.781	4.910	193.0E6	185.2E6	73.775	74.894
12)	B 4,4'-DDE	5.959	5.101	172.6E6	157.6E6	73.946	76.358
13)	MA Dieldrin	6.104	5.231	183.0E6	176.3E6	73.687	75.442
14)	MA Endrin	6.331	5.505	161.0E6	157.9E6	74.982	75.307
15)	B Endosulfa...	6.553	5.801	149.9E6	141.0E6	73.985	74.848
16)	A 4,4'-DDD	6.474	5.655	137.0E6	121.7E6	73.449	75.283
17)	MA 4,4'-DDT	6.785	5.904	151.5E6	135.8E6	73.903	75.365
18)	B Endrin al...	6.682	5.981	120.0E6	115.4E6	74.014	74.842
19)	B Endosulfa...	6.917	6.204	152.9E6	146.2E6	73.611	75.038
20)	A Methoxychlor	7.266	6.484	78299274	74169350	73.430	74.972
21)	B Endrin ke...	7.396	6.709	156.0E6	163.2E6	73.695	75.183
22)	Mirex	7.861	6.881	118.6E6	139.2E6	71.918	74.425

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032223\
Data File : PL081606.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Mar 2023 18:29
Operator : AR\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 00:50:15 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
Quant Title : GC Extractables
QLast Update : Thu Mar 23 00:50:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

