

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032319\
 Data File : PL046373.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2019 19:01
 Operator : AJ\SJ
 Sample : K2014-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PST-026-SW037-031819

Manual Integrations
 APPROVED

Sohil
 3/25/2019 11:31:59 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 25 04:35:13 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.360	3.978	140.2E6	42524199	19.275	18.573
28)	SA Decachlor...	8.039	9.034	119.3E6	35855869	14.178	15.465
Target Compounds							
12)	B 4,4'-DDE	5.478	6.314	35417519	8587522	3.847	3.459
17)	MA 4,4'-DDT	6.224	7.102	15433251	4421646	2.038	2.109m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032319\
Data File : PL046373.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2019 19:01
Operator : AJ\SJ
Sample : K2014-03
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PST-026-SW037-031819


Manual Integrations
APPROVED

Sohil
3/25/2019 11:31:59 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 25 04:35:13 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.360 min
Delta R.T.: -0.001 min
Response: 140217151
Conc: 19.27 ng/ml

Instrument :
ECD_L
ClientSampleId :
PST-026-SW037-031819


Manual Integrations
APPROVED

Sohil
3/25/2019 11:31:59 AM




#1 Tetrachloro-m-xylene

R.T.: 3.978 min
Delta R.T.: -0.002 min
Response: 42524199
Conc: 18.57 ng/ml




#12 4,4'-DDE

R.T.: 5.478 min
Delta R.T.: -0.003 min
Response: 35417519
Conc: 3.85 ng/ml



#12 4,4'-DDE

R.T.: 6.314 min
Delta R.T.: -0.004 min
Response: 8587522
Conc: 3.46 ng/ml




#17 4,4'-DDT

R.T.: 6.224 min
Delta R.T.: -0.004 min
Response: 15433251
Conc: 2.04 ng/ml

Instrument :
ECD_L
ClientSampleId :
PST-026-SW037-031819


Manual Integrations
APPROVED

Sohil
3/25/2019 11:31:59 AM




#17 4,4'-DDT

R.T.: 7.102 min
Delta R.T.: -0.005 min
Response: 4421646
Conc: 2.11 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.039 min
Delta R.T.: -0.005 min
Response: 119250969
Conc: 14.18 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.034 min
Delta R.T.: -0.006 min
Response: 35855869
Conc: 15.47 ng/ml