

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032321\
 Data File : PL065494.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2021 13:18
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 3/24/2021 11:22:13 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 04:12:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 19 02:14:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.286	3.785	129.7E6	130.9E6	54.631	47.677
2)	SA Decachlor...	7.925	8.733	118.1E6	107.1E6	48.870	54.031
Target Compounds							
2)	A alpha-BHC	3.709	4.170	202.5E6	186.6E6	54.155	45.907
3)	MA gamma-BHC...	3.983	4.448	181.8E6	168.6E6	53.149	46.170m
4)	MA Heptachlor	4.268	4.935	168.7E6	152.7E6	51.536	45.786
5)	MB Aldrin	4.506	5.231	166.7E6	146.6E6	52.613	44.752
6)	B beta-BHC	4.233	4.631	75418072	75486369	53.333	43.818
7)	B delta-BHC	4.425	4.838	177.6E6	160.2E6	53.385m	53.618
8)	B Heptachlo...	4.938	5.617	152.2E6	135.5E6	51.456	46.871
9)	A Endosulfan I	5.266	5.968	138.8E6	120.7E6	50.916	45.846
10)	B gamma-Chl...	5.159	5.848	150.3E6	130.9E6	50.915	46.897
11)	B alpha-Chl...	5.215	5.918	148.5E6	128.8E6	51.035	46.395
12)	B 4,4'-DDE	5.380	6.082	133.3E6	113.6E6	50.660m	45.604
13)	MA Dieldrin	5.501	6.223	140.9E6	121.5E6	50.344m	44.667
14)	MA Endrin	5.751	6.438	124.9E6	104.5E6	50.014m	47.296
15)	B Endosulfa...	6.025	6.655	116.5E6	102.8E6	47.564	45.537
16)	A 4,4'-DDD	5.888	6.572	108.9E6	94629422	49.289	43.023
17)	MA 4,4'-DDT	6.119	6.865	100.6E6	93325758	53.168m	47.548
18)	B Endrin al...	6.191	6.780	87403284	79520704	45.516	45.392
19)	B Endosulfa...	6.401	7.001	120.3E6	104.4E6	48.693	47.572
20)	A Methoxychlor	6.665	7.331	58157049	51770914	50.725m	49.572
21)	B Endrin ke...	6.879	7.466	127.2E6	111.5E6	47.330	47.246m
22)	Mirex	7.052	7.902	96471551	83326173	47.376	50.062

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032321\
Data File : PL065494.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2021 13:18
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
3/24/2021 11:22:13 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 04:12:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
Quant Title : GC Extractables
QLast Update : Fri Mar 19 02:14:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

