

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032321\
 Data File : PL065510.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2021 18:05
 Operator : AR\AJ
 Sample : M1765-04
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RBR-54

Manual Integrations
 APPROVED

Ankita
 3/24/2021 11:22:30 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 04:16:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 19 02:14:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.286	3.784	54083535	56156982	22.774	20.460
28)	SA Decachlor...	7.924	8.733	32421420	28351471	13.419	14.304
Target Compounds							
6)	B beta-BHC	4.241	4.634	4239109	3020320	2.998m	1.753 #
12)	B 4,4'-DDE	5.380	6.080	5331080	3592747	2.026m	1.442 #
17)	MA 4,4'-DDT	6.120	6.864	4544268	3566045	2.401	1.817

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032321\
Data File : PL065510.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2021 18:05
Operator : AR\AJ
Sample : M1765-04
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampled :
RBR-54


Manual Integrations
APPROVED

Ankita
3/24/2021 11:22:30 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 04:16:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
Quant Title : GC Extractables
QLast Update : Fri Mar 19 02:14:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.286 min
Delta R.T.: -0.003 min
Response: 54083535
Conc: 22.77 ng/ml

Instrument :
ECD_L
ClientSampleId :
RBR-54


Manual Integrations
APPROVED

Ankita
3/24/2021 11:22:30 AM




#1 Tetrachloro-m-xylene

R.T.: 3.784 min
Delta R.T.: -0.006 min
Response: 56156982
Conc: 20.46 ng/ml




#6 beta-BHC

R.T.: 4.241 min
Delta R.T.: 0.003 min
Response: 4239109
Conc: 3.00 ng/ml m



#6 beta-BHC

R.T.: 4.634 min
Delta R.T.: -0.003 min
Response: 3020320
Conc: 1.75 ng/ml




#12 4,4'-DDE

R.T.: 5.380 min
Delta R.T.: -0.008 min
Response: 5331080
Conc: 2.03 ng/ml m

Instrument :
ECD_L
ClientSampleId :
RBR-54


Manual Integrations
APPROVED

Ankita
3/24/2021 11:22:30 AM




#12 4,4'-DDE

R.T.: 6.080 min
Delta R.T.: -0.009 min
Response: 3592747
Conc: 1.44 ng/ml




#17 4,4'-DDT

R.T.: 6.120 min
Delta R.T.: -0.007 min
Response: 4544268
Conc: 2.40 ng/ml



#17 4,4'-DDT

R.T.: 6.864 min
Delta R.T.: -0.009 min
Response: 3566045
Conc: 1.82 ng/ml




#28 Decachlorobiphenyl

R.T.: 7.924 min
Delta R.T.: -0.009 min
Response: 32421420
Conc: 13.42 ng/ml

Instrument :
ECD_L
ClientSampleId :
RBR-54

Manual Integrations
APPROVED

Ankita
3/24/2021 11:22:30 AM



#28 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: -0.010 min
Response: 28351471
Conc: 14.30 ng/ml