

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032323\
 Data File : PL081672.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Mar 2023 13:03
 Operator : AR\AJ
 Sample : 02023-05MS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-13-TEST-PIT-ON-GRAVELLYMS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 03/24/2023
 Supervised By :Ankita Jodhani 03/24/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 24 03:03:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 23 01:07:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.328	2.658	43935315	36856025	18.890	18.464
28) SA Decachlor...	8.750	7.776	34578188	35969418	19.321	18.137
Target Compounds						
2) A alpha-BHC	3.778	3.157	183.0E6	145.2E6	53.386	49.805
3) MA gamma-BHC...	4.104	3.485	173.6E6	144.7E6	54.648m	51.959
4) MA Heptachlor	4.685	3.818	160.1E6	140.2E6	50.790m	48.888
5) MB Aldrin	5.022	4.096	156.3E6	138.9E6	54.636	52.122
6) B beta-BHC	4.307	3.784	76206665	59264210	50.633m	49.761
7) B delta-BHC	4.550	4.011	165.5E6	133.4E6	54.470	51.579
8) B Heptachlo...	5.448	4.598	141.5E6	123.2E6	52.036	50.721
9) A Endosulfan I	5.830	4.966	123.8E6	116.4E6	50.085	52.745
10) B gamma-Chl...	5.703	4.847	152.5E6	137.3E6	58.086	56.474
11) B alpha-Chl...	5.782	4.910	148.6E6	133.2E6	57.425	54.823
12) B 4,4'-DDE	5.959	5.101	122.8E6	104.6E6	54.502	53.098
13) MA Dieldrin	6.104	5.230	130.9E6	123.3E6	54.059	54.815
14) MA Endrin	6.331	5.505	110.3E6	113.1E6	52.151	55.887
15) B Endosulfa...	6.553	5.800	106.7E6	103.5E6	50.258	57.057
16) A 4,4'-DDD	6.474	5.655	97829086	83813914	54.244	55.230
17) MA 4,4'-DDT	6.786	5.904	102.6E6	93150201	50.554	54.054
18) B Endrin al...	6.683	5.981	90312410	84203586	55.199	54.220
19) B Endosulfa...	6.917	6.204	107.0E6	100.4E6	51.328	51.596
20) A Methoxychlor	7.266	6.483	54649648	49971424	50.748	50.170
21) B Endrin ke...	7.396	6.708	114.1E6	112.8E6	54.481	52.783
22) Mirex	7.860	6.881	85154933	92479804	49.703m	48.173

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032232\
Data File : PL081672.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Mar 2023 13:03
Operator : AR\AJ
Sample : 02023-05MS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-13-TEST-PIT-ON-GRAVELLYMS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 03/24/2023
Supervised By :Ankita Jodhani 03/24/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 24 03:03:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
Quant Title : GC Extractables
QLast Update : Thu Mar 23 01:07:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

