

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\pl032423\
 Data File : PL081697.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Mar 2023 11:20
 Operator : AR\AJ
 Sample : PB151631BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB151631BS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/27/2023
 Supervised By : Ankita Jodhani 03/27/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 21:19:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 23 01:07:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.330	2.658	38304244	41881417	16.469	20.981 #
28) SA Decachlor...	8.752	7.777	31006002	35629027	17.325	17.965
Target Compounds						
23) Chlordane-1	4.474	3.646	23115310	17174088	232.505	222.968
24) Chlordane-2	5.000	4.220	25175354	20948323	241.894	224.243
25) Chlordane-3	5.704	4.848	92730008	56865548	234.100m	225.773
26) Chlordane-4	5.786	4.909	113.0E6	59191972	236.061	225.982
27) Chlordane-5	6.628	5.575	18668190	15622644	219.843	234.080

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\pl032423\
Data File : PL081697.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Mar 2023 11:20
Operator : AR\AJ
Sample : PB151631BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB151631BS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/27/2023
Supervised By : Ankita Jodhani 03/27/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 21:19:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
Quant Title : GC Extractables
QLast Update : Thu Mar 23 01:07:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

