

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032519\
 Data File : PL046461.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Mar 2019 20:33
 Operator : AJ\SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 01:17:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.359	3.978	145.5E6	47802865	20.007	20.878
28)	SA Decachlor...	8.037	9.031	169.2E6	50494993	20.117	21.779
Target Compounds							
8)	B Heptachlo...	0.000	5.842	0	1872397	N.D.	0.710 #
10)	B gamma-Chl...	0.000	6.090	0	1609432	N.D.	0.567 #
25)	Chlordane-3	0.000	6.090	0	1609432	N.D.	4.221 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032519\
Data File : PL046461.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Mar 2019 20:33
Operator : AJ\SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 01:17:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.359 min
Delta R.T.: -0.002 min
Response: 145544463
Conc: 20.01 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.978 min
Delta R.T.: -0.003 min
Response: 47802865
Conc: 20.88 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 5.034 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 5.842 min
Delta R.T.: -0.011 min
Response: 1872397
Conc: 0.71 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T. : 5.257 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
I.BLK




#10 gamma-Chlordane

R.T.: 6.090 min
Delta R.T.: 0.002 min
Response: 1609432
Conc: 0.57 ng/ml




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T. : 5.258 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 6.090 min
Delta R.T.: 0.000 min
Response: 1609432
Conc: 4.22 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.037 min
Delta R.T.: -0.007 min
Response: 169206094
Conc: 20.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK



#28 Decachlorobiphenyl

R.T.: 9.031 min
Delta R.T.: -0.009 min
Response: 50494993
Conc: 21.78 ng/ml