

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032521\
 Data File : PL065571.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Mar 2021 17:31
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 3/26/2021 11:07:11 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 02:20:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 19 02:14:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.285	3.784	143.0E6	149.6E6	60.235	54.507
28)	SA Decachlor...	7.923	8.732	113.2E6	108.0E6	46.868	54.477
Target Compounds							
2)	A alpha-BHC	3.709	4.169	223.6E6	216.2E6	59.797	53.207
3)	MA gamma-BHC...	3.980	4.447	200.3E6	196.8E6	58.563m	53.906m
4)	MA Heptachlor	4.267	4.934	176.2E6	174.6E6	53.843	52.367
5)	MB Aldrin	4.504	5.230	180.1E6	170.5E6	56.841	52.073
6)	B beta-BHC	4.232	4.631	80225056	87750882	56.732	50.937
7)	B delta-BHC	4.424	4.838	193.6E6	176.7E6	58.197m	59.130
8)	B Heptachlo...	4.937	5.615	161.8E6	159.0E6	54.723	54.990m
9)	A Endosulfan I	5.264	5.966	146.9E6	141.6E6	53.906	53.797
10)	B gamma-Chl...	5.158	5.847	160.5E6	159.0E6	54.365	56.964
11)	B alpha-Chl...	5.214	5.917	157.7E6	151.4E6	54.197	54.553
12)	B 4,4'-DDE	5.380	6.081	145.3E6	131.1E6	55.203	52.637
13)	MA Dieldrin	5.501	6.222	148.3E6	138.8E6	52.961	51.021
14)	MA Endrin	5.749	6.438	131.1E6	115.3E6	52.493m	52.222
15)	B Endosulfa...	6.023	6.654	120.2E6	113.9E6	49.063	50.474
16)	A 4,4'-DDD	5.884	6.572	126.4E6	121.1E6	57.178m	55.061
17)	MA 4,4'-DDT	6.118	6.865	89338893	87926343	47.201m	44.797
18)	B Endrin al...	6.190	6.778	90773161	88439736	47.271	50.483
19)	B Endosulfa...	6.399	7.000	124.3E6	114.2E6	50.311	52.049
20)	A Methoxychlor	6.663	7.330	52433225	48551024	45.733m	46.489
21)	B Endrin ke...	6.876	7.466	133.9E6	124.1E6	49.808m	52.618
22)	Mirex	7.050	7.902	95286288	87674085	46.794	52.674

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032521\
Data File : PL065571.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Mar 2021 17:31
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
3/26/2021 11:07:11 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 02:20:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
Quant Title : GC Extractables
QLast Update : Fri Mar 19 02:14:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

