

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032620\
 Data File : PL057610.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Mar 2020 23:50
 Operator : AJ\MA
 Sample : L2034-01MSD
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 FK-BPM-04-008-ABCDMSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 01:27:08 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031420.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 16 09:40:37 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.458	3.957	217.0E6	51203417	23.415	23.601
28)	SA Decachlor...	8.207	8.981	199.4E6	46528228	18.845	20.984
Target Compounds							
2)	A alpha-BHC	3.895	4.345	912.3E6	188.2E6	57.270	58.587
3)	MA gamma-BHC...	4.179	4.627	823.7E6	182.3E6	57.306	58.101
4)	MA Heptachlor	4.479	5.129	853.2E6	193.5E6	57.448	61.476
5)	MB Aldrin	4.727	5.431	800.5E6	189.4E6	58.033	62.095
6)	B beta-BHC	4.431	4.800	352.6E6	89010055	59.981	61.712
7)	B delta-BHC	4.635	5.014	797.5E6	167.0E6	56.819	63.194
8)	B Heptachlo...	5.170	5.816	732.0E6	178.3E6	57.113	65.346
9)	A Endosulfan I	5.508	6.170	680.0E6	173.7E6	57.193	61.056
10)	B gamma-Chl...	5.397	6.049	742.6E6	191.3E6	56.946	62.079
11)	B alpha-Chl...	5.455	6.121	716.0E6	187.2E6	55.407	60.152
12)	B 4,4'-DDE	5.619	6.280	718.9E6	175.6E6	57.714	61.767
13)	MA Dieldrin	5.751	6.426	728.7E6	180.5E6	57.773	61.027
14)	MA Endrin	6.008	6.642	607.1E6	138.9E6	53.111	57.581
15)	B Endosulfa...	6.281	6.853	584.3E6	141.2E6	53.802	53.760
16)	A 4,4'-DDD	6.135	6.768	558.0E6	134.8E6	56.935	59.047
17)	MA 4,4'-DDT	6.374	7.065	536.3E6	121.6E6	53.773	53.393
18)	B Endrin al...	6.450	6.976	498.3E6	123.1E6	53.955	55.375
19)	B Endosulfa...	6.663	7.199	554.4E6	136.3E6	55.207	55.879
20)	A Methoxychlor	6.920	7.525	258.4E6	66557039	48.152	51.175
21)	B Endrin ke...	7.153	7.667	609.3E6	144.4E6	53.909	55.794
22)	Mirex	7.341	8.117	420.0E6	102.7E6	51.861	53.807

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032620\
Data File : PL057610.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Mar 2020 23:50
Operator : AJ\MA
Sample : L2034-01MSD
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
FK-BPM-04-008-ABCDMSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 01:27:08 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031420.M
Quant Title : GC Extractables
QLast Update : Mon Mar 16 09:40:37 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

