

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032719\
 Data File : PL046543.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2019 14:32
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 01:11:47 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.359	3.978	346.3E6	108.3E6	47.609	47.310
2)	SA Decachlor...	8.035	9.030	349.1E6	108.0E6	41.502	46.572
Target Compounds							
2)	A alpha-BHC	3.788	4.367	571.5E6	153.8E6	49.006	47.872
3)	MA gamma-BHC...	4.065	4.651	518.9E6	142.5E6	48.291	47.615
4)	MA Heptachlor	4.352	5.158	529.4E6	145.6E6	48.013	47.349
5)	MB Aldrin	4.592	5.462	482.6E6	133.2E6	46.787	46.488
6)	B beta-BHC	4.318	4.824	210.7E6	63458865	46.790	47.868
7)	B delta-BHC	4.513	5.040	484.7E6	118.1E6	47.886	50.897
8)	B Heptachlo...	5.029	5.848	446.7E6	123.0E6	45.794	46.659
9)	A Endosulfan I	5.361	6.204	415.3E6	119.5E6	44.168	45.802
10)	B gamma-Chl...	5.252	6.082	457.7E6	129.7E6	44.558	45.700
11)	B alpha-Chl...	5.309	6.154	447.9E6	129.6E6	44.176	45.679
12)	B 4,4'-DDE	5.475	6.312	406.6E6	113.2E6	44.165	45.613
13)	MA Dieldrin	5.600	6.460	436.5E6	121.8E6	43.495	44.927
14)	MA Endrin	5.853	6.677	413.2E6	107.4E6	43.864	44.953
15)	B Endosulfa...	6.128	6.888	401.1E6	109.3E6	44.771	45.283
16)	A 4,4'-DDD	5.986	6.803	335.5E6	91629966	44.386	45.541
17)	MA 4,4'-DDT	6.222	7.100	360.7E6	96070006	47.626	45.814
18)	B Endrin al...	6.295	7.012	331.8E6	94488809	44.271	44.486
19)	B Endosulfa...	6.506	7.234	363.0E6	106.6E6	44.206	45.205
20)	A Methoxychlor	6.766	7.558	185.5E6	58488712	43.654	45.882
21)	B Endrin ke...	6.990	7.704	392.5E6	109.1E6	42.448	43.919
22)	Mirex	7.166	8.156	296.0E6	90280942	40.627	45.137

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032719\
Data File : PL046543.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2019 14:32
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 01:11:47 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

