

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032819\
 Data File : PL046690.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Mar 2019 05:16
 Operator : AJ\SJ
 Sample : K2107-02MS
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 29 05:28:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.977	138.1E6	45079928	18.980	19.689
28)	SA Decachlor...	8.035	9.029	192.7E6	54263211	22.908	23.405
Target Compounds							
2)	A alpha-BHC	3.788	4.367	654.9E6	186.5E6	56.159	58.032
3)	MA gamma-BHC...	4.065	4.651	607.3E6	174.9E6	56.518	58.427
4)	MA Heptachlor	4.352	5.157	633.9E6	179.4E6	57.495	58.337
5)	MB Aldrin	4.592	5.461	612.1E6	172.4E6	59.340	60.178
6)	B beta-BHC	4.318	4.824	247.0E6	76724653	54.869	57.875
7)	B delta-BHC	4.513	5.039	576.0E6	146.6E6	56.898	63.182
8)	B Heptachlo...	5.029	5.847	567.0E6	157.8E6	58.129	59.842
9)	A Endosulfan I	5.361	6.203	536.3E6	151.0E6	57.039	57.865
10)	B gamma-Chl...	5.252	6.081	584.0E6	165.0E6	56.853	58.144
11)	B alpha-Chl...	5.308	6.154	570.7E6	163.3E6	56.287	57.564
12)	B 4,4'-DDE	5.475	6.312	542.2E6	147.6E6	58.886	59.444
13)	MA Dieldrin	5.601	6.460	584.2E6	158.6E6	58.216	58.517
14)	MA Endrin	5.854	6.678	519.3E6	133.1E6	55.129	55.693
15)	B Endosulfa...	6.128	6.888	506.1E6	140.1E6	56.492	58.067
16)	A 4,4'-DDD	5.987	6.802	457.2E6	123.9E6	60.489	61.576
17)	MA 4,4'-DDT	6.221	7.100	429.3E6	119.1E6	56.679	56.794
18)	B Endrin al...	6.295	7.012	425.1E6	121.3E6	56.719	57.095
19)	B Endosulfa...	6.506	7.235	491.7E6	139.4E6	59.877	59.094
20)	A Methoxychlor	6.766	7.558	225.6E6	68694277	53.082	53.887
21)	B Endrin ke...	6.989	7.704	566.6E6	153.0E6	61.273	61.566
22)	Mirex	7.165	8.156	402.1E6	116.4E6	55.188	58.209

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032819\
Data File : PL046690.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Mar 2019 05:16
Operator : AJ\SJ
Sample : K2107-02MS
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 29 05:28:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

