

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032919\
 Data File : PL046782.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2019 03:49
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 05:48:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.977	369.2E6	117.4E6	50.745	51.261
28)	SA Decachlor...	8.034	9.028	409.6E6	119.6E6	48.697	51.594
Target Compounds							
2)	A alpha-BHC	3.787	4.366	602.6E6	171.7E6	51.671	53.438
3)	MA gamma-BHC...	4.064	4.650	552.5E6	159.6E6	51.424	53.319
4)	MA Heptachlor	4.351	5.156	555.2E6	153.5E6	50.358	49.892
5)	MB Aldrin	4.591	5.461	539.2E6	153.5E6	52.272	53.576
6)	B beta-BHC	4.317	4.823	225.3E6	71037956	50.032	53.586
7)	B delta-BHC	4.512	5.039	534.1E6	142.6E6	52.764	61.459
8)	B Heptachlor...	5.028	5.847	501.9E6	157.8E6	51.456	59.849
9)	A Endosulfan I	5.360	6.203	480.0E6	140.7E6	51.050	53.897
10)	B gamma-Chl...	5.251	6.081	527.5E6	170.1E6	51.351	59.925
11)	B alpha-Chl...	5.307	6.153	511.6E6	154.0E6	50.458	54.275
12)	B 4,4'-DDE	5.474	6.311	482.4E6	129.5E6	52.396	52.154
13)	MA Dieldrin	5.599	6.460	515.9E6	140.3E6	51.411	51.750
14)	MA Endrin	5.853	6.677	451.3E6	121.0E6	47.912	50.614
15)	B Endosulfa...	6.127	6.887	453.3E6	127.5E6	50.595	52.821
16)	A 4,4'-DDD	5.986	6.801	415.5E6	113.1E6	54.963	56.220
17)	MA 4,4'-DDT	6.220	7.099	352.4E6	98690934	46.521	47.064
18)	B Endrin al...	6.294	7.010	375.3E6	109.8E6	50.067	51.709
19)	B Endosulfa...	6.505	7.234	417.6E6	120.0E6	50.862	50.870
20)	A Methoxychlor	6.764	7.557	202.5E6	60898891	47.658	47.772
21)	B Endrin ke...	6.988	7.703	479.0E6	134.2E6	51.802	54.020
22)	Mirex	7.164	8.155	350.3E6	103.2E6	48.086	51.598
24)	Chlordane-2	0.000	5.461f	0	153.5E6	N.D.	1314.062 #
25)	Chlordane-3	5.251	6.081	527.5E6	170.1E6	473.663	446.068
26)	Chlordane-4	5.307	6.153	511.6E6	154.0E6	494.749	333.665 #
27)	Chlordane-5	6.127	0.000	453.3E6	0	1328.325	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032919\
Data File : PL046782.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Mar 2019 03:49
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 05:48:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.358 min
Delta R.T.: -0.003 min
Response: 369153342
Conc: 50.75 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.977 min
Delta R.T.: -0.004 min
Response: 117367323
Conc: 51.26 ng/ml




#2 alpha-BHC

R.T.: 3.787 min
Delta R.T.: -0.004 min
Response: 602592002
Conc: 51.67 ng/ml



#2 alpha-BHC


R.T.: 4.366 min
Delta R.T.: -0.005 min
Response: 171712854
Conc: 53.44 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.064 min
Delta R.T.: -0.004 min
Response: 552523070
Conc: 51.42 ng/ml

Instrument :
ECD_L
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 4.650 min
Delta R.T.: -0.005 min
Response: 159586118
Conc: 53.32 ng/ml




#4 Heptachlor

R.T.: 4.351 min
Delta R.T.: -0.005 min
Response: 555209610
Conc: 50.36 ng/ml



#4 Heptachlor


R.T.: 5.156 min
Delta R.T.: -0.006 min
Response: 153462548
Conc: 49.89 ng/ml



#5 Aldrin


R.T.: 4.591 min
Delta R.T.: -0.005 min
Response: 539182203
Conc: 52.27 ng/ml

Instrument :
ECD_L
ClientSampleId :




#5 Aldrin

R.T.: 5.461 min
Delta R.T.: -0.006 min
Response: 153494587
Conc: 53.58 ng/ml




#6 beta-BHC

R.T.: 4.317 min
Delta R.T.: -0.004 min
Response: 225266829
Conc: 50.03 ng/ml



#6 beta-BHC


R.T.: 4.823 min
Delta R.T.: -0.004 min
Response: 71037956
Conc: 53.59 ng/ml



#7 delta-BHC


R.T.: 4.512 min
Delta R.T.: -0.005 min
Response: 534136961
Conc: 52.76 ng/ml

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 5.039 min
Delta R.T.: -0.005 min
Response: 142594980
Conc: 61.46 ng/ml




#8 Heptachlor epoxide

R.T.: 5.028 min
Delta R.T.: -0.005 min
Response: 501874344
Conc: 51.46 ng/ml



#8 Heptachlor epoxide


R.T.: 5.847 min
Delta R.T.: -0.007 min
Response: 157832897
Conc: 59.85 ng/ml



#9 Endosulfan I


R.T.: 5.360 min
Delta R.T.: -0.006 min
Response: 479987996
Conc: 51.05 ng/ml

Instrument :
ECD_L
ClientSampleId :




#9 Endosulfan I

R.T.: 6.203 min
Delta R.T.: -0.007 min
Response: 140663995
Conc: 53.90 ng/ml




#10 gamma-Chlordane

R.T.: 5.251 min
Delta R.T.: -0.006 min
Response: 527506304
Conc: 51.35 ng/ml



#10 gamma-Chlordane


R.T.: 6.081 min
Delta R.T.: -0.007 min
Response: 170071161
Conc: 59.92 ng/ml



#11 alpha-Chlordane


R.T.: 5.307 min
Delta R.T.: -0.007 min
Response: 511600306
Conc: 50.46 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 alpha-Chlordane

R.T.: 6.153 min
Delta R.T.: -0.007 min
Response: 153961186
Conc: 54.28 ng/ml




#12 4,4'-DDE

R.T.: 5.474 min
Delta R.T.: -0.007 min
Response: 482410277
Conc: 52.40 ng/ml



#12 4,4'-DDE


R.T.: 6.311 min
Delta R.T.: -0.007 min
Response: 129464306
Conc: 52.15 ng/ml



#13 Dieldrin


R.T.: 5.599 min
Delta R.T.: -0.007 min
Response: 515939736
Conc: 51.41 ng/ml

Instrument :
ECD_L
ClientSampleId :




#13 Dieldrin

R.T.: 6.460 min
Delta R.T.: -0.007 min
Response: 140276372
Conc: 51.75 ng/ml




#14 Endrin

R.T.: 5.853 min
Delta R.T.: -0.007 min
Response: 451295294
Conc: 47.91 ng/ml



#14 Endrin


R.T.: 6.677 min
Delta R.T.: -0.008 min
Response: 120956556
Conc: 50.61 ng/ml



#15 Endosulfan II


R.T.: 6.127 min
Delta R.T.: -0.007 min
Response: 453313428
Conc: 50.60 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 6.887 min
Delta R.T.: -0.008 min
Response: 127450591
Conc: 52.82 ng/ml




#16 4,4'-DDD

R.T.: 5.986 min
Delta R.T.: -0.007 min
Response: 415465031
Conc: 54.96 ng/ml



#16 4,4'-DDD


R.T.: 6.801 min
Delta R.T.: -0.008 min
Response: 113116795
Conc: 56.22 ng/ml



#17 4,4'-DDT


R.T.: 6.220 min
Delta R.T.: -0.007 min
Response: 352368273
Conc: 46.52 ng/ml

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 7.099 min
Delta R.T.: -0.008 min
Response: 98690934
Conc: 47.06 ng/ml




#18 Endrin aldehyde

R.T.: 6.294 min
Delta R.T.: -0.008 min
Response: 375259701
Conc: 50.07 ng/ml



#18 Endrin aldehyde


R.T.: 7.010 min
Delta R.T.: -0.008 min
Response: 109830356
Conc: 51.71 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.505 min
Delta R.T.: -0.008 min
Response: 417637947
Conc: 50.86 ng/ml

Instrument :
ECD_L
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 7.234 min
Delta R.T.: -0.008 min
Response: 120002746
Conc: 50.87 ng/ml




#20 Methoxychlor

R.T.: 6.764 min
Delta R.T.: -0.009 min
Response: 202542186
Conc: 47.66 ng/ml



#20 Methoxychlor


R.T.: 7.557 min
Delta R.T.: -0.008 min
Response: 60898891
Conc: 47.77 ng/ml



#21 Endrin ketone


R.T.: 6.988 min
Delta R.T.: -0.008 min
Response: 478998372
Conc: 51.80 ng/ml

Instrument :
ECD_L
ClientSampleId :




#21 Endrin ketone

R.T.: 7.703 min
Delta R.T.: -0.009 min
Response: 134214500
Conc: 54.02 ng/ml




#22 Mirex

R.T.: 7.164 min
Delta R.T.: -0.009 min
Response: 350329491
Conc: 48.09 ng/ml



#22 Mirex


R.T.: 8.155 min
Delta R.T.: -0.010 min
Response: 103204625
Conc: 51.60 ng/ml



#24 Chlordane-2


R.T.: 0.000 min
Exp R.T.: 4.701 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#24 Chlordane-2

R.T.: 5.461 min
Delta R.T.: 0.018 min
Response: 153494587
Conc: 1314.06 ng/ml




#25 Chlordane-3

R.T.: 5.251 min
Delta R.T.: -0.008 min
Response: 527506304
Conc: 473.66 ng/ml



#25 Chlordane-3


R.T.: 6.081 min
Delta R.T.: -0.008 min
Response: 170071161
Conc: 446.07 ng/ml



#26 Chlordane-4


R.T.: 5.307 min
Delta R.T.: -0.007 min
Response: 511600306
Conc: 494.75 ng/ml

Instrument :
ECD_L
ClientSampleId :




#26 Chlordane-4

R.T.: 6.153 min
Delta R.T.: -0.010 min
Response: 153961186
Conc: 333.66 ng/ml




#27 Chlordane-5

R.T.: 6.127 min
Delta R.T.: -0.012 min
Response: 453313428
Conc: 1328.32 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T. : 6.967 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.034 min
Delta R.T.: -0.010 min
Response: 409589833
Conc: 48.70 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.028 min
Delta R.T.: -0.011 min
Response: 119618825
Conc: 51.59 ng/ml