

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032919\
 Data File : PL046727.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Mar 2019 14:38
 Operator : AJ\SJ
 Sample : PB118397BS
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB118397BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 29 23:13:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.359	3.978	124.8E6	40806004	17.162	17.822
2)	SA Decachlor...	8.038	9.032	159.6E6	48531811	18.974	20.933
Target Compounds							
2)	A alpha-BHC	3.789	4.368	589.4E6	163.2E6	50.538	50.794
3)	MA gamma-BHC...	4.066	4.653	539.6E6	151.5E6	50.221	50.625
4)	MA Heptachlor	4.353	5.159	556.7E6	155.7E6	50.492	50.634
5)	MB Aldrin	4.593	5.463	542.4E6	151.2E6	52.583	52.760
6)	B beta-BHC	4.319	4.825	220.5E6	67553020	48.963	50.957
7)	B delta-BHC	4.515	5.041	502.9E6	127.6E6	49.678	55.004
8)	B Heptachlo...	5.031	5.849	502.9E6	137.6E6	51.565	52.164
9)	A Endosulfan I	5.363	6.206	477.6E6	133.8E6	50.792	51.251
10)	B gamma-Chl...	5.254	6.083	516.1E6	145.4E6	50.242	51.219
11)	B alpha-Chl...	5.309	6.156	505.3E6	144.0E6	49.833	50.768
12)	B 4,4'-DDE	5.477	6.314	477.5E6	129.8E6	51.865	52.302
13)	MA Dieldrin	5.602	6.463	513.8E6	139.3E6	51.203	51.405
14)	MA Endrin	5.855	6.680	428.0E6	110.3E6	45.434	46.151
15)	B Endosulfa...	6.130	6.890	445.0E6	123.0E6	49.672	50.961
16)	A 4,4'-DDD	5.988	6.804	404.2E6	109.9E6	53.467	54.603
17)	MA 4,4'-DDT	6.223	7.102	365.1E6	100.6E6	48.198	47.959
18)	B Endrin al...	6.297	7.014	379.8E6	108.3E6	50.675	50.971
19)	B Endosulfa...	6.508	7.237	427.1E6	122.2E6	52.017	51.821
20)	A Methoxychlor	6.768	7.561	193.8E6	59112399	45.598	46.371
21)	B Endrin ke...	6.991	7.707	505.6E6	135.8E6	54.675	54.645
22)	Mirex	7.167	8.159	351.5E6	103.1E6	48.241	51.550

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032919\
Data File : PL046727.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Mar 2019 14:38
Operator : AJ\SJ
Sample : PB118397BS
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB118397BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 29 23:13:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

