

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032919\
 Data File : PL046762.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Mar 2019 23:02
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 3/30/2019 10:11:51 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 02:17:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.359	3.977	362.9E6	118.4E6	49.892	51.731
28)	SA Decachlor...	8.034	9.028	417.4E6	120.4E6	49.629	51.926
Target Compounds							
2)	A alpha-BHC	3.788	4.367	597.9E6	171.9E6	51.269	53.508
3)	MA gamma-BHC...	4.065	4.651	550.3E6	160.0E6	51.217	53.467
4)	MA Heptachlor	4.352	5.157	557.5E6	159.3E6	50.564	51.804
5)	MB Aldrin	4.592	5.461	537.4E6	154.3E6	52.101	53.852
6)	B beta-BHC	4.318	4.824	225.1E6	71055587	49.988	53.599
7)	B delta-BHC	4.513	5.038	532.4E6	142.5E6	52.589	61.437m
8)	B Heptachlo...	5.029	5.846	502.3E6	152.9E6	51.498	57.974m
9)	A Endosulfan I	5.360	6.204	479.3E6	142.1E6	50.982	54.440
10)	B gamma-Chl...	5.252	6.080	526.3E6	151.0E6	51.229	53.196m
11)	B alpha-Chl...	5.308	6.152	511.3E6	151.8E6	50.427	53.518m
12)	B 4,4'-DDE	5.475	6.312	495.7E6	131.6E6	53.838	53.029
13)	MA Dieldrin	5.600	6.460	519.7E6	142.7E6	51.789	52.651
14)	MA Endrin	5.853	6.677	445.2E6	119.9E6	47.265	50.155
15)	B Endosulfa...	6.127	6.887	458.1E6	128.3E6	51.131	53.168
16)	A 4,4'-DDD	5.986	6.802	418.8E6	114.8E6	55.405	57.040
17)	MA 4,4'-DDT	6.221	7.099	360.1E6	99640076	47.546	47.517
18)	B Endrin al...	6.295	7.011	382.0E6	111.4E6	50.969	52.433
19)	B Endosulfa...	6.505	7.234	421.6E6	120.8E6	51.345	51.217
20)	A Methoxychlor	6.765	7.558	201.7E6	60759352	47.460	47.663
21)	B Endrin ke...	6.987	7.703	517.9E6	136.7E6	56.014m	55.020
22)	Mirex	7.164	8.155	352.0E6	104.2E6	48.319	52.118

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032919\
Data File : PL046762.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Mar 2019 23:02
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
3/30/2019 10:11:51 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 02:17:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

