

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032919\
 Data File : PL046779.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2019 03:06
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 3/30/2019 10:11:58 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 03:34:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.977	372.0E6	121.1E6	51.134	52.869
28)	SA Decachlor...	8.035	9.029	412.4E6	120.0E6	49.028	51.778
Target Compounds							
2)	A alpha-BHC	3.788	4.367	608.2E6	175.7E6	52.156	54.677
3)	MA gamma-BHC...	4.065	4.651	558.1E6	161.7E6	51.943	54.021
4)	MA Heptachlor	4.352	5.158	554.9E6	156.9E6	50.329	51.025
5)	MB Aldrin	4.592	5.461	540.6E6	154.5E6	52.414	53.929
6)	B beta-BHC	4.318	4.824	227.0E6	72032557	50.415	54.336
7)	B delta-BHC	4.514	5.040	538.2E6	142.7E6	53.168	61.483
8)	B Heptachlo...	5.029	5.846	502.4E6	153.8E6	51.510	58.308m
9)	A Endosulfan I	5.361	6.203	477.4E6	141.2E6	50.770	54.099
10)	B gamma-Chl...	5.252	6.080	526.0E6	150.6E6	51.209	53.079m
11)	B alpha-Chl...	5.308	6.152	509.8E6	148.0E6	50.281	52.175m
12)	B 4,4'-DDE	5.474	6.312	482.2E6	130.2E6	52.373m	52.446
13)	MA Dieldrin	5.601	6.460	518.1E6	142.9E6	51.630	52.706
14)	MA Endrin	5.853	6.676	452.1E6	111.8E6	48.003	46.800m
15)	B Endosulfa...	6.128	6.888	454.7E6	128.4E6	50.751	53.206
16)	A 4,4'-DDD	5.987	6.802	417.9E6	115.0E6	55.289	57.156
17)	MA 4,4'-DDT	6.222	7.100	346.5E6	99602300	45.742	47.499
18)	B Endrin al...	6.295	7.012	376.6E6	110.6E6	50.247	52.079
19)	B Endosulfa...	6.506	7.234	418.1E6	120.8E6	50.914	51.193
20)	A Methoxychlor	6.766	7.558	200.1E6	60687492	47.075	47.606
21)	B Endrin ke...	6.988	7.704	514.2E6	135.5E6	55.605m	54.548
22)	Mirex	7.166	8.155	351.6E6	103.9E6	48.264	51.962

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032919\
Data File : PL046779.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Mar 2019 03:06
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
3/30/2019 10:11:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 03:34:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

