

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032921\
 Data File : PL065647.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Mar 2021 16:36
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 29 17:29:44 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 19 02:14:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.283	3.783	128.2E6	130.8E6	53.996	47.655
28)	SA Decachlor...	7.919	8.731	105.1E6	99649566	43.506	50.274
Target Compounds							
2)	A alpha-BHC	3.706	4.168	203.1E6	194.6E6	54.311	47.885
3)	MA gamma-BHC...	3.979	4.447	183.6E6	175.7E6	53.672	48.104
4)	MA Heptachlor	4.264	4.932	168.1E6	160.4E6	51.361	48.112
5)	MB Aldrin	4.501	5.229	170.3E6	155.1E6	53.738	47.354
6)	B beta-BHC	4.230	4.630	76338094	80605586	53.983	46.789
7)	B delta-BHC	4.423	4.837	179.7E6	166.4E6	54.038	55.665
8)	B Heptachlo...	4.933	5.615	152.5E6	145.8E6	51.566	50.429
9)	A Endosulfan I	5.260	5.965	142.3E6	130.8E6	52.201	49.697
10)	B gamma-Chl...	5.154	5.846	153.2E6	141.4E6	51.891	50.678
11)	B alpha-Chl...	5.210	5.915	150.1E6	139.1E6	51.586	50.115
12)	B 4,4'-DDE	5.376	6.080	135.2E6	123.6E6	51.386	49.636
13)	MA Dieldrin	5.497	6.221	143.4E6	132.8E6	51.227	48.814
14)	MA Endrin	5.746	6.436	116.0E6	103.5E6	46.466	46.850
15)	B Endosulfa...	6.019	6.653	120.1E6	110.3E6	49.043	48.851
16)	A 4,4'-DDD	5.882	6.571	122.6E6	112.7E6	55.474	51.248
17)	MA 4,4'-DDT	6.116	6.863	84757003	86663291	44.780	44.153
18)	B Endrin al...	6.186	6.777	91114444	88633664	47.449	50.593
19)	B Endosulfa...	6.396	6.999	125.4E6	113.8E6	50.744	51.860
20)	A Methoxychlor	6.661	7.329	49584001	49030989	43.248	46.948
21)	B Endrin ke...	6.874	7.465	138.5E6	128.1E6	51.509	54.297
22)	Mirex	7.046	7.900	99486342	88024167	48.856	52.884

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL032921\
Data File : PL065647.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Mar 2021 16:36
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 29 17:29:44 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031821.M
Quant Title : GC Extractables
QLast Update : Fri Mar 19 02:14:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

