

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL033020\
 Data File : PL057678.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2020 11:58
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 31 01:23:53 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031420.M
 Quant Title : GC Extractables
 QLast Update : Mon Mar 16 09:40:37 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.460	3.958	502.0E6	106.9E6	54.154	49.287
2)	SA Decachlor...	8.207	8.979	450.9E6	94491457	42.616	42.614
Target Compounds							
2)	A alpha-BHC	3.897	4.345	872.4E6	163.4E6	54.768	50.856
3)	MA gamma-BHC...	4.180	4.627	806.9E6	162.1E6	56.139	51.658
4)	MA Heptachlor	4.481	5.129	810.0E6	167.3E6	54.539	53.165
5)	MB Aldrin	4.728	5.431	758.1E6	166.4E6	54.954	54.558
6)	B beta-BHC	4.432	4.800	318.9E6	75207697	54.241	52.143
7)	B delta-BHC	4.636	5.014	778.6E6	153.5E6	55.478	58.389
8)	B Heptachlo...	5.171	5.816	683.8E6	158.6E6	53.353	58.132
9)	A Endosulfan I	5.510	6.170	613.9E6	148.0E6	51.629	52.011
10)	B gamma-Chl...	5.398	6.049	684.1E6	166.6E6	52.465	54.065
11)	B alpha-Chl...	5.456	6.121	672.6E6	166.1E6	52.047	53.357
12)	B 4,4'-DDE	5.620	6.280	648.7E6	149.9E6	52.083	52.723
13)	MA Dieldrin	5.753	6.426	658.1E6	153.5E6	52.180	51.886
14)	MA Endrin	6.010	6.642	569.4E6	119.0E6	49.814	49.318
15)	B Endosulfa...	6.282	6.852	525.5E6	121.5E6	48.383	46.238
16)	A 4,4'-DDD	6.136	6.768	494.7E6	111.0E6	50.474	48.585
17)	MA 4,4'-DDT	6.376	7.065	486.8E6	104.3E6	48.815	45.801
18)	B Endrin al...	6.451	6.976	440.8E6	103.6E6	47.731	46.599
19)	B Endosulfa...	6.664	7.199	473.8E6	112.0E6	47.179	45.934
20)	A Methoxychlor	6.921	7.525	241.2E6	56077696	44.945	43.117
21)	B Endrin ke...	7.154	7.667	537.1E6	114.3E6	47.519	44.163
22)	Mirex	7.341	8.116	361.9E6	84409503	44.684	44.204

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL033020\
Data File : PL057678.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Mar 2020 11:58
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 31 01:23:53 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031420.M
Quant Title : GC Extractables
QLast Update : Mon Mar 16 09:40:37 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

