

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL033023\
 Data File : PL081782.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Mar 2023 09:20
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 30 21:09:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 23 01:07:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.335	2.654	120.3E6	97358132	51.724	48.773
28) SA Decachlor...	8.761	7.775	84907467	100.3E6	47.444	50.548
Target Compounds						
2) A alpha-BHC	3.785	3.153	182.0E6	144.9E6	53.095	49.713
3) MA gamma-BHC...	4.114	3.481	169.6E6	136.4E6	53.374	48.992
4) MA Heptachlor	4.693	3.815	163.3E6	145.1E6	51.806	50.620
5) MB Aldrin	5.030	4.093	147.4E6	134.3E6	51.538	50.414
6) B beta-BHC	4.317	3.781	73297666	57327238	48.700	48.134
7) B delta-BHC	4.558	4.008	162.8E6	129.2E6	53.576	49.955
8) B Heptachlo...	5.456	4.595	139.3E6	124.9E6	51.198	51.415
9) A Endosulfan I	5.837	4.963	123.5E6	113.2E6	49.995	51.290
10) B gamma-Chl...	5.711	4.844	134.7E6	125.6E6	51.315	51.689
11) B alpha-Chl...	5.789	4.907	132.1E6	125.6E6	51.038	51.693
12) B 4,4'-DDE	5.967	5.098	113.3E6	104.2E6	50.284	52.897
13) MA Dieldrin	6.111	5.227	123.5E6	117.3E6	51.030	52.146
14) MA Endrin	6.339	5.502	105.9E6	107.0E6	50.061	52.888
15) B Endosulfa...	6.560	5.798	99040275	94730127	46.671	52.220
16) A 4,4'-DDD	6.482	5.653	87581659	82264354	48.562	54.209
17) MA 4,4'-DDT	6.794	5.902	96553332	87061703	47.572	50.521
18) B Endrin al...	6.690	5.979	82813889	78288942	50.616	50.412
19) B Endosulfa...	6.924	6.202	103.7E6	99413774	49.736	51.094
20) A Methoxychlor	7.274	6.482	51785577	50536351	48.088	50.737
21) B Endrin ke...	7.404	6.707	106.4E6	110.4E6	50.787	51.648
22) Mirex	7.869	6.879	84901403	100.4E6	49.555	52.325

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL033023\
Data File : PL081782.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Mar 2023 09:20
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 30 21:09:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032223.M
Quant Title : GC Extractables
QLast Update : Thu Mar 23 01:07:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

