

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL033121\
 Data File : PL065740.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Mar 2021 21:42
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 02:45:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 31 15:57:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.280	3.782	147.4E6	150.0E6	56.564	54.649
28)	SA Decachlor...	7.914	8.729	116.2E6	104.7E6	50.171	46.448
Target Compounds							
2)	A alpha-BHC	3.703	4.167	238.3E6	238.6E6	58.261	58.722
3)	MA gamma-BHC...	3.976	4.446	213.0E6	205.9E6	57.535	56.244
4)	MA Heptachlor	4.261	4.931	195.2E6	190.7E6	54.190	54.037
5)	MB Aldrin	4.498	5.228	201.1E6	185.4E6	58.895	56.165
6)	B beta-BHC	4.227	4.629	81823067	94841210	51.335	54.450
7)	B delta-BHC	4.420	4.836	202.5E6	179.4E6	56.286	60.145
8)	B Heptachlo...	4.930	5.614	180.1E6	171.9E6	57.278	55.343
9)	A Endosulfan I	5.256	5.964	154.5E6	146.8E6	53.620	53.489
10)	B gamma-Chl...	5.150	5.845	175.1E6	169.6E6	55.639	56.518
11)	B alpha-Chl...	5.206	5.914	167.6E6	156.8E6	53.866	52.999
12)	B 4,4'-DDE	5.373	6.079	166.3E6	166.8E6	60.268	63.842
13)	MA Dieldrin	5.493	6.221	159.4E6	164.7E6	54.562	58.434
14)	MA Endrin	5.743	6.435	140.7E6	135.2E6	52.117	54.289
15)	B Endosulfa...	6.015	6.652	131.8E6	151.0E6	53.088	63.911
16)	A 4,4'-DDD	5.878	6.570	136.2E6	126.3E6	60.046	59.927
17)	MA 4,4'-DDT	6.111	6.863	136.4E6	126.2E6	64.747	56.381
18)	B Endrin al...	6.182	6.777	116.6E6	96574767	61.616	50.896
19)	B Endosulfa...	6.392	6.998	157.1E6	127.1E6	60.903	52.378
20)	A Methoxychlor	6.656	7.328	58703192	90547752	46.761	71.493 #
21)	B Endrin ke...	6.869	7.464	144.3E6	137.3E6	52.603	53.412
22)	Mirex	7.041	7.899	96173684	88967574	45.923	45.582
23)	Chlordane-1	0.000	4.736	0	2828467	N.D.	25.609 #
24)	Chlordane-2	0.000	5.228f	0	185.4E6	N.D.	1620.446 #
25)	Chlordane-3	5.150	5.845	175.1E6	169.6E6	563.340	398.953 #
26)	Chlordane-4	5.206	5.914	167.6E6	156.8E6	587.026	310.010 #
27)	Chlordane-5	6.015	0.000	131.8E6	0	1289.032	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL033121\
 Data File : PL065740.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Mar 2021 21:42
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 02:45:04 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 31 15:57:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.280 min
Delta R.T.: 0.000 min
Response: 147350573
Conc: 56.56 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.782 min
Delta R.T.: 0.000 min
Response: 150008664
Conc: 54.65 ng/ml




#2 alpha-BHC

R.T.: 3.703 min
Delta R.T.: 0.000 min
Response: 238280821
Conc: 58.26 ng/ml



#2 alpha-BHC


R.T.: 4.167 min
Delta R.T.: 0.000 min
Response: 238588703
Conc: 58.72 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 3.976 min
Delta R.T.: 0.000 min
Response: 212975733
Conc: 57.53 ng/ml

Instrument :
ECD_L
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 4.446 min
Delta R.T.: 0.000 min
Response: 205919284
Conc: 56.24 ng/ml




#4 Heptachlor

R.T.: 4.261 min
Delta R.T.: 0.000 min
Response: 195213202
Conc: 54.19 ng/ml



#4 Heptachlor


R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 190709993
Conc: 54.04 ng/ml



#5 Aldrin


R.T.: 4.498 min
Delta R.T.: 0.000 min
Response: 201079226
Conc: 58.90 ng/ml

Instrument :
ECD_L
ClientSampleId :




#5 Aldrin

R.T.: 5.228 min
Delta R.T.: 0.000 min
Response: 185441106
Conc: 56.17 ng/ml




#6 beta-BHC

R.T.: 4.227 min
Delta R.T.: 0.000 min
Response: 81823067
Conc: 51.34 ng/ml



#6 beta-BHC


R.T.: 4.629 min
Delta R.T.: 0.000 min
Response: 94841210
Conc: 54.45 ng/ml



#7 delta-BHC


R.T.: 4.420 min
Delta R.T.: 0.000 min
Response: 202478155
Conc: 56.29 ng/ml

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 4.836 min
Delta R.T.: 0.000 min
Response: 179382961
Conc: 60.14 ng/ml




#8 Heptachlor epoxide

R.T.: 4.930 min
Delta R.T.: 0.000 min
Response: 180110264
Conc: 57.28 ng/ml



#8 Heptachlor epoxide


R.T.: 5.614 min
Delta R.T.: 0.000 min
Response: 171859557
Conc: 55.34 ng/ml



#9 Endosulfan I


R.T.: 5.256 min
Delta R.T.: 0.000 min
Response: 154515126
Conc: 53.62 ng/ml

Instrument :
ECD_L
ClientSampleId :




#9 Endosulfan I

R.T.: 5.964 min
Delta R.T.: 0.000 min
Response: 146845110
Conc: 53.49 ng/ml




#10 gamma-Chlordane

R.T.: 5.150 min
Delta R.T.: 0.000 min
Response: 175091816
Conc: 55.64 ng/ml



#10 gamma-Chlordane


R.T.: 5.845 min
Delta R.T.: 0.000 min
Response: 169571912
Conc: 56.52 ng/ml



#11 alpha-Chlordane


R.T.: 5.206 min
Delta R.T.: 0.000 min
Response: 167603837
Conc: 53.87 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 alpha-Chlordane

R.T.: 5.914 min
Delta R.T.: 0.000 min
Response: 156828278
Conc: 53.00 ng/ml




#12 4,4'-DDE

R.T.: 5.373 min
Delta R.T.: 0.000 min
Response: 166281105
Conc: 60.27 ng/ml



#12 4,4'-DDE


R.T.: 6.079 min
Delta R.T.: 0.001 min
Response: 166788671
Conc: 63.84 ng/ml



#13 Dieldrin


R.T.: 5.493 min
Delta R.T.: 0.000 min
Response: 159408518
Conc: 54.56 ng/ml

Instrument :
ECD_L
ClientSampleId :




#13 Dieldrin

R.T.: 6.221 min
Delta R.T.: 0.001 min
Response: 164684497
Conc: 58.43 ng/ml




#14 Endrin

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 140716056
Conc: 52.12 ng/ml



#14 Endrin


R.T.: 6.435 min
Delta R.T.: 0.001 min
Response: 135220373
Conc: 54.29 ng/ml



#15 Endosulfan II


R.T.: 6.015 min
Delta R.T.: 0.000 min
Response: 131845924
Conc: 53.09 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 6.652 min
Delta R.T.: 0.001 min
Response: 151048402
Conc: 63.91 ng/ml




#16 4,4'-DDD

R.T.: 5.878 min
Delta R.T.: 0.000 min
Response: 136246304
Conc: 60.05 ng/ml



#16 4,4'-DDD


R.T.: 6.570 min
Delta R.T.: 0.000 min
Response: 126309196
Conc: 59.93 ng/ml



#17 4,4'-DDT


R.T.: 6.111 min
Delta R.T.: 0.000 min
Response: 136410280
Conc: 64.75 ng/ml

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 6.863 min
Delta R.T.: 0.002 min
Response: 126240671
Conc: 56.38 ng/ml




#18 Endrin aldehyde

R.T.: 6.182 min
Delta R.T.: 0.000 min
Response: 116585605
Conc: 61.62 ng/ml



#18 Endrin aldehyde


R.T.: 6.777 min
Delta R.T.: 0.001 min
Response: 96574767
Conc: 50.90 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.392 min
Delta R.T.: 0.000 min
Response: 157145798
Conc: 60.90 ng/ml

Instrument :
ECD_L
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 6.998 min
Delta R.T.: 0.001 min
Response: 127065425
Conc: 52.38 ng/ml




#20 Methoxychlor

R.T.: 6.656 min
Delta R.T.: 0.000 min
Response: 58703192
Conc: 46.76 ng/ml



#20 Methoxychlor


R.T.: 7.328 min
Delta R.T.: 0.002 min
Response: 90547752
Conc: 71.49 ng/ml



#21 Endrin ketone


R.T.: 6.869 min
Delta R.T.: 0.000 min
Response: 144257954
Conc: 52.60 ng/ml

Instrument :
ECD_L
ClientSampleId :




#21 Endrin ketone

R.T.: 7.464 min
Delta R.T.: 0.002 min
Response: 137284562
Conc: 53.41 ng/ml




#22 Mirex

R.T.: 7.041 min
Delta R.T.: 0.000 min
Response: 96173684
Conc: 45.92 ng/ml



#22 Mirex


R.T.: 7.899 min
Delta R.T.: 0.002 min
Response: 88967574
Conc: 45.58 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 4.114 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#23 Chlordane-1

R.T.: 4.736 min
Delta R.T.: -0.009 min
Response: 2828467
Conc: 25.61 ng/ml




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T.: 4.600 min
Response: 0
Conc: N.D.



#24 Chlordane-2


R.T.: 5.228 min
Delta R.T.: 0.020 min
Response: 185441106
Conc: 1620.45 ng/ml



#25 Chlordane-3


R.T.: 5.150 min
Delta R.T.: 0.000 min
Response: 175091816
Conc: 563.34 ng/ml

Instrument :
ECD_L
ClientSampleId :




#25 Chlordane-3

R.T.: 5.845 min
Delta R.T.: 0.000 min
Response: 169571912
Conc: 398.95 ng/ml




#26 Chlordane-4

R.T.: 5.206 min
Delta R.T.: 0.000 min
Response: 167603837
Conc: 587.03 ng/ml



#26 Chlordane-4


R.T.: 5.914 min
Delta R.T.: 0.000 min
Response: 156828278
Conc: 310.01 ng/ml



#27 Chlordane-5


R.T.: 6.015 min
Delta R.T.: -0.002 min
Response: 131845924
Conc: 1289.03 ng/ml

Instrument :
ECD_L
ClientSampleId :




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.712 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 7.914 min
Delta R.T.: 0.000 min
Response: 116198079
Conc: 50.17 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.729 min
Delta R.T.: 0.000 min
Response: 104718612
Conc: 46.45 ng/ml