

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL033121\
 Data File : PL065740.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Mar 2021 21:42
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 4/1/2021 12:58:02 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 01 03:01:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 31 15:57:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.280	3.782	147.4E6	150.0E6	56.564	54.649
28)	SA Decachlor...	7.914	8.729	116.2E6	104.7E6	50.171	46.448
Target Compounds							
2)	A alpha-BHC	3.703	4.167	238.3E6	238.6E6	58.261	58.722
3)	MA gamma-BHC...	3.976	4.446	213.0E6	205.9E6	57.535	56.244
4)	MA Heptachlor	4.261	4.931	195.2E6	190.7E6	54.190	54.037
5)	MB Aldrin	4.498	5.228	201.1E6	185.4E6	58.895	56.165
6)	B beta-BHC	4.227	4.629	81823067	94841210	51.335	54.450
7)	B delta-BHC	4.420	4.835	202.5E6	178.2E6	56.286	59.754m
8)	B Heptachlo...	4.930	5.614	180.1E6	171.9E6	57.278	55.343
9)	A Endosulfan I	5.256	5.964	154.5E6	146.8E6	53.620	53.489
10)	B gamma-Chl...	5.150	5.845	175.1E6	169.6E6	55.639	56.518
11)	B alpha-Chl...	5.206	5.914	167.6E6	156.8E6	53.866	52.999
12)	B 4,4'-DDE	5.371	6.078	163.9E6	147.9E6	59.408m	56.617m
13)	MA Dieldrin	5.493	6.221	159.4E6	164.7E6	54.562	58.434
14)	MA Endrin	5.743	6.435	140.7E6	135.2E6	52.117	54.289
15)	B Endosulfa...	6.015	6.651	131.8E6	132.2E6	53.088	55.936m
16)	A 4,4'-DDD	5.876	6.570	135.9E6	126.3E6	59.877m	59.927
17)	MA 4,4'-DDT	6.110	6.863	115.5E6	126.2E6	54.831m	56.381
18)	B Endrin al...	6.181	6.777	107.0E6	96574767	56.551m	50.896
19)	B Endosulfa...	6.390	6.998	133.4E6	127.1E6	51.716m	52.378
20)	A Methoxychlor	6.656	7.327	58703192	61699846	46.761	48.716m
21)	B Endrin ke...	6.869	7.464	144.3E6	137.3E6	52.603	53.412
22)	Mirex	7.041	7.899	96173684	88967574	45.923	45.582

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL033121\
Data File : PL065740.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Mar 2021 21:42
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
4/1/2021 12:58:02 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 01 03:01:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
Quant Title : GC Extractables
QLast Update : Wed Mar 31 15:57:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

