

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040119\
 Data File : PL046809.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2019 16:34
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 4/2/2019 10:26:20 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 01:25:25 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.978	335.8E6	104.2E6	46.158	45.524
28)	SA Decachlor...	8.035	9.031	417.9E6	120.0E6	49.684	51.764
Target Compounds							
2)	A alpha-BHC	3.788	4.368	552.6E6	148.4E6	47.381	46.170
3)	MA gamma-BHC...	4.065	4.652	505.1E6	138.8E6	47.010	46.389
4)	MA Heptachlor	4.352	5.158	526.1E6	143.0E6	47.722	46.493
5)	MB Aldrin	4.592	5.462	495.0E6	135.0E6	47.990	47.118
6)	B beta-BHC	4.318	4.825	204.3E6	62039942	45.383	46.798
7)	B delta-BHC	4.513	5.040	482.1E6	118.3E6	47.621	50.974
8)	B Heptachlor...	5.029	5.848	470.9E6	128.8E6	48.278	48.821
9)	A Endosulfan I	5.361	6.205	446.5E6	122.9E6	47.485	47.109
10)	B gamma-Chl...	5.252	6.082	486.3E6	133.3E6	47.339	46.983
11)	B alpha-Chl...	5.308	6.155	479.7E6	133.4E6	47.311	47.028
12)	B 4,4'-DDE	5.474	6.313	448.3E6	117.9E6	48.697m	47.513
13)	MA Dieldrin	5.601	6.461	480.6E6	127.6E6	47.891	47.079
14)	MA Endrin	5.854	6.679	448.0E6	112.9E6	47.558	47.252
15)	B Endosulfa...	6.128	6.889	428.9E6	115.9E6	47.871	48.041
16)	A 4,4'-DDD	5.987	6.803	362.4E6	96829748	47.945	48.125
17)	MA 4,4'-DDT	6.222	7.100	379.3E6	101.8E6	50.076	48.543
18)	B Endrin al...	6.295	7.012	355.6E6	100.6E6	47.448	47.365
19)	B Endosulfa...	6.507	7.235	400.4E6	112.2E6	48.765	47.580
20)	A Methoxychlor	6.767	7.559	204.7E6	62251660	48.163	48.833
21)	B Endrin ke...	6.989	7.705	452.9E6	120.8E6	48.979	48.601
22)	Mirex	7.167	8.157	345.8E6	98632263	47.469	49.312

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040119\
 Data File : PL046809.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2019 16:34
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 4/2/2019 10:26:20 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 01:25:25 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

