

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040124\
 Data File : PL088842.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2024 10:02
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/02/2024
 Supervised By : Ankita Jodhani 04/02/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 01:12:23 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031324.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 13 15:45:55 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.643	2.853	133.1E6	71063125	50.221	53.321
28)	SA Decachlor...	9.312	8.082	115.0E6	70956073	46.520	52.331
Target Compounds							
2)	A alpha-BHC	4.108	3.367	205.2E6	107.5E6	51.415	53.496
3)	MA gamma-BHC...	4.446	3.704	196.7E6	104.2E6	50.724	54.274
4)	MA Heptachlor	5.046	4.053	189.2E6	99024407	49.818	55.022
5)	MB Aldrin	5.394	4.338	186.2E6	101.0E6	50.803	55.865
6)	B beta-BHC	4.644	4.007	83404633	44464189	47.030	51.869
7)	B delta-BHC	4.896	4.241	193.5E6	102.3E6	48.038	53.452
8)	B Heptachlo...	5.825	4.849	172.0E6	90587757	50.075	55.085
9)	A Endosulfan I	6.215	5.225	153.5E6	83403613	50.236	55.482
10)	B gamma-Chl...	6.084	5.104	170.6E6	92565853	51.222	55.887
11)	B alpha-Chl...	6.164	5.168	168.8E6	92457190	50.767	55.699
12)	B 4,4'-DDE	6.339	5.361	151.1E6	84167146	50.564	57.705
13)	MA Dieldrin	6.494	5.494	164.1E6	91100653	50.622	56.091
14)	MA Endrin	6.726	5.773	134.5E6	78469170	49.201	54.438
15)	B Endosulfa...	6.946	6.071	143.4E6	78663776	50.052	54.708
16)	A 4,4'-DDD	6.861	5.923	127.4E6	71549231	50.023	57.452
17)	MA 4,4'-DDT	7.179	6.178	128.5E6	74033085	48.218	55.158
18)	B Endrin al...	7.076	6.253	107.6E6	60405857	51.558m	57.642
19)	B Endosulfa...	7.315	6.478	134.5E6	76010452	48.723	54.785
20)	A Methoxychlor	7.659	6.761	69219663	41453517	45.947	51.683
21)	B Endrin ke...	7.806	6.991	145.9E6	90893962	49.366	55.064
22)	Mirex	8.288	7.177	112.6E6	73459914	48.012	54.478

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040124\
Data File : PL088842.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Apr 2024 10:02
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 04/02/2024

Supervised By :Ankita Jodhani 04/02/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 01:12:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031324.M
Quant Title : GC Extractables
QLast Update : Wed Mar 13 15:45:55 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

