

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040125\
 Data File : PL094982.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Apr 2025 11:26
 Operator : AR\AJ
 Sample : PEM
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PEM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 02 01:22:48 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
 Quant Title : GC Extractables
 QLast Update : Tue Mar 11 17:42:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.536	2.769	64919749	84567061	22.934	23.693
28)	SA Decachlor...	9.053	7.903	54432393	100.5E6	25.828	24.884
Target Compounds							
2)	A alpha-BHC	3.992	3.271	48957171	61527431	11.790	11.412
3)	MA gamma-BHC...	4.325	3.601	46060191	58534736	11.543	11.389
6)	B beta-BHC	4.524	3.902	21717113	28116469	11.769	12.658
14)	MA Endrin	6.573	5.630	134.9E6	235.1E6	48.669	53.876
16)	A 4,4'-DDD	6.708	5.777	7316668	9697013	3.378	2.697
17)	MA 4,4'-DDT	7.023	6.027	260.8E6	495.4E6	109.628	122.873
18)	B Endrin al...	0.000	6.104	0	8788581	N.D.	2.611 #
20)	A Methoxychlor	7.499	6.603	318.1E6	593.3E6	265.714	279.720
21)	B Endrin ke...	7.642	6.831	10727917	28355999	4.059	5.941 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040125\
Data File : PL094982.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Apr 2025 11:26
Operator : AR\AJ
Sample : PEM
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PEM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 02 01:22:48 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031125.M
Quant Title : GC Extractables
QLast Update : Tue Mar 11 17:42:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.536 min
Delta R.T.: -0.002 min
Response: 64919749
Conc: 22.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#1 Tetrachloro-m-xylene

R.T.: 2.769 min
Delta R.T.: -0.002 min
Response: 84567061
Conc: 23.69 ng/ml




#2 alpha-BHC

R.T.: 3.992 min
Delta R.T.: -0.002 min
Response: 48957171
Conc: 11.79 ng/ml



#2 alpha-BHC


R.T.: 3.271 min
Delta R.T.: -0.003 min
Response: 61527431
Conc: 11.41 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.325 min
Delta R.T.: -0.002 min
Response: 46060191
Conc: 11.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#3 gamma-BHC (Lindane)

R.T.: 3.601 min
Delta R.T.: -0.003 min
Response: 58534736
Conc: 11.39 ng/ml




#6 beta-BHC

R.T.: 4.524 min
Delta R.T.: -0.001 min
Response: 21717113
Conc: 11.77 ng/ml



#6 beta-BHC


R.T.: 3.902 min
Delta R.T.: -0.003 min
Response: 28116469
Conc: 12.66 ng/ml



#14 Endrin


R.T.: 6.573 min
Delta R.T.: -0.001 min
Response: 134913113
Conc: 48.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#14 Endrin

R.T.: 5.630 min
Delta R.T.: -0.003 min
Response: 235094397
Conc: 53.88 ng/ml




#16 4,4'-DDD

R.T.: 6.708 min
Delta R.T.: -0.002 min
Response: 7316668
Conc: 3.38 ng/ml



#16 4,4'-DDD


R.T.: 5.777 min
Delta R.T.: -0.005 min
Response: 9697013
Conc: 2.70 ng/ml



#17 4,4'-DDT


R.T.: 7.023 min
Delta R.T.: -0.001 min
Response: 260754742
Conc: 109.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#17 4,4'-DDT

R.T.: 6.027 min
Delta R.T.: -0.004 min
Response: 495436119
Conc: 122.87 ng/ml




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T. : 6.924 min
Response: 0
Conc: N.D.



#18 Endrin aldehyde


R.T.: 6.104 min
Delta R.T.: -0.004 min
Response: 8788581
Conc: 2.61 ng/ml



#20 Methoxychlor


R.T.: 7.499 min
Delta R.T.: 0.000 min
Response: 318078267
Conc: 265.71 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM




#20 Methoxychlor

R.T.: 6.603 min
Delta R.T.: -0.004 min
Response: 593305873
Conc: 279.72 ng/ml




#21 Endrin ketone

R.T.: 7.642 min
Delta R.T.: -0.002 min
Response: 10727917
Conc: 4.06 ng/ml



#21 Endrin ketone


R.T.: 6.831 min
Delta R.T.: -0.004 min
Response: 28355999
Conc: 5.94 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.053 min
Delta R.T.: -0.003 min
Response: 54432393
Conc: 25.83 ng/ml

Instrument :
ECD_L
ClientSampleId :
PEM



#28 Decachlorobiphenyl

R.T.: 7.903 min
Delta R.T.: -0.004 min
Response: 100516426
Conc: 24.88 ng/ml