

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040219\
 Data File : PL046921.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Apr 2019 20:58
 Operator : AJ\SJ
 Sample : K2056-05MS
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 CS-7MS

Manual Integrations
 APPROVED

Sohil
 4/3/2019 5:56:38 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 01:21:21 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.977	134.9E6	44330113	18.549	19.361
28)	SA Decachlor...	8.035	9.029	174.7E6	51113190	20.773	22.046
Target Compounds							
2)	A alpha-BHC	3.787	4.367	630.1E6	174.8E6	54.034	54.409
3)	MA gamma-BHC...	4.064	4.651	583.0E6	165.1E6	54.258	55.159
4)	MA Heptachlor	4.351	5.157	617.4E6	172.7E6	55.997	56.136
5)	MB Aldrin	4.591	5.461	584.3E6	163.5E6	56.650	57.057
6)	B beta-BHC	4.318	4.824	235.6E6	73100468	52.316	55.141
7)	B delta-BHC	4.513	5.040	538.6E6	135.6E6	53.208	58.465
8)	B Heptachlor...	5.028	5.847	551.6E6	150.0E6	56.550	56.889
9)	A Endosulfan I	5.360	6.203	403.4E6	113.9E6	42.906	43.654
10)	B gamma-Chl...	5.251	6.082	566.1E6	158.8E6	55.107	55.944
11)	B alpha-Chl...	5.307	6.154	554.9E6	158.4E6	54.732	55.839
12)	B 4,4'-DDE	5.475	6.312	515.6E6	140.6E6	55.998	56.623
13)	MA Dieldrin	5.600	6.460	562.7E6	151.7E6	56.069	55.980
14)	MA Endrin	5.852	6.677	530.5E6	134.8E6	56.324	56.411
15)	B Endosulfa...	6.127	6.888	297.6E6	82506071	33.216	34.194
16)	A 4,4'-DDD	5.984	6.802	478.4E6	114.2E6	63.293m	56.765
17)	MA 4,4'-DDT	6.221	7.100	452.8E6	121.6E6	59.787	57.991
18)	B Endrin al...	6.294	7.012	402.9E6	114.7E6	53.757	53.979
19)	B Endosulfa...	6.506	7.234	476.9E6	132.7E6	58.080	56.253
20)	A Methoxychlor	6.765	7.558	238.1E6	70334354	56.031	55.174
21)	B Endrin ke...	6.988	7.703	522.0E6	138.2E6	56.455	55.625
22)	Mirex	7.164	8.156	383.6E6	107.6E6	52.657	53.785

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040219\
Data File : PL046921.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Apr 2019 20:58
Operator : AJ\SJ
Sample : K2056-05MS
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
CS-7MS

Manual Integrations
APPROVED

Sohil
4/3/2019 5:56:38 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 01:21:21 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

