

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040219\
 Data File : PL046939.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2019 01:17
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleID :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 4/3/2019 5:56:48 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 05:52:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.977	373.0E6	118.4E6	51.273	51.711
28)	SA Decachlor...	8.035	9.030	427.8E6	121.3E6	50.856	52.331
Target Compounds							
2)	A alpha-BHC	3.787	4.367	627.6E6	171.9E6	53.816	53.489
3)	MA gamma-BHC...	4.064	4.651	568.4E6	160.0E6	52.902	53.465
4)	MA Heptachlor	4.351	5.156	594.0E6	166.7E6	53.873	54.207m
5)	MB Aldrin	4.591	5.461	555.6E6	154.0E6	53.864	53.763
6)	B beta-BHC	4.317	4.824	230.5E6	70912937	51.189	53.491
7)	B delta-BHC	4.513	5.040	550.1E6	140.9E6	54.340	60.720
8)	B Heptachlor...	5.028	5.846	521.4E6	148.1E6	53.461	56.176m
9)	A Endosulfan I	5.360	6.203	497.2E6	139.3E6	52.879	53.387
10)	B gamma-Chl...	5.251	6.080	545.8E6	150.4E6	53.136	52.993m
11)	B alpha-Chl...	5.307	6.154	534.5E6	151.9E6	52.719	53.542
12)	B 4,4'-DDE	5.474	6.312	505.1E6	131.0E6	54.862m	52.762
13)	MA Dieldrin	5.600	6.460	537.5E6	142.5E6	53.561	52.585
14)	MA Endrin	5.853	6.677	505.2E6	128.5E6	53.640	53.758
15)	B Endosulfa...	6.127	6.888	475.2E6	128.8E6	53.040	53.375
16)	A 4,4'-DDD	5.986	6.802	407.6E6	107.3E6	53.926	53.312
17)	MA 4,4'-DDT	6.221	7.100	427.1E6	113.1E6	56.388	53.943
18)	B Endrin al...	6.294	7.012	391.8E6	110.9E6	52.269	52.231
19)	B Endosulfa...	6.505	7.235	439.2E6	123.2E6	53.492	52.214
20)	A Methoxychlor	6.765	7.559	226.7E6	67179029	53.335	52.699
21)	B Endrin ke...	6.987	7.704	503.9E6	131.1E6	54.492m	52.751
22)	Mirex	7.165	8.156	375.8E6	104.2E6	51.583	52.089

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040219\
Data File : PL046939.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2019 01:17
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
4/3/2019 5:56:48 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 05:52:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

