

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040219\
 Data File : PL046942.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2019 02:00
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 4/3/2019 5:56:52 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 05:33:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.977	373.8E6	119.8E6	51.390	52.313
28)	SA Decachlor...	8.035	9.029	381.0E6	115.0E6	45.302	49.606
Target Compounds							
2)	A alpha-BHC	3.787	4.367	620.6E6	171.7E6	53.216	53.441
3)	MA gamma-BHC...	4.064	4.651	542.4E6	157.8E6	50.480	52.720
4)	MA Heptachlor	4.351	5.157	570.9E6	161.5E6	51.785	52.504
5)	MB Aldrin	4.591	5.461	533.1E6	150.7E6	51.681	52.596
6)	B beta-BHC	4.317	4.824	222.6E6	70186089	49.435	52.943
7)	B delta-BHC	4.513	5.040	531.2E6	137.9E6	52.478	59.441
8)	B Heptachlor...	5.028	5.846	502.9E6	144.6E6	51.558	54.818m
9)	A Endosulfan I	5.360	6.204	481.1E6	133.4E6	51.164	51.123
10)	B gamma-Chl...	5.251	6.082	529.2E6	145.8E6	51.512	51.389
11)	B alpha-Chl...	5.308	6.154	518.6E6	144.6E6	51.145	50.971
12)	B 4,4'-DDE	5.473	6.312	479.1E6	126.1E6	52.037m	50.799
13)	MA Dieldrin	5.600	6.460	510.9E6	136.1E6	50.910	50.205
14)	MA Endrin	5.853	6.677	465.8E6	120.6E6	49.450	50.453
15)	B Endosulfa...	6.128	6.888	433.5E6	122.8E6	48.380	50.884
16)	A 4,4'-DDD	5.986	6.802	377.9E6	102.7E6	49.989	51.038
17)	MA 4,4'-DDT	6.221	7.099	382.9E6	108.3E6	50.549	51.669
18)	B Endrin al...	6.295	7.011	356.6E6	105.4E6	47.580	49.614
19)	B Endosulfa...	6.506	7.234	398.4E6	115.8E6	48.520	49.109
20)	A Methoxychlor	6.766	7.558	201.5E6	63583590	47.410	49.878
21)	B Endrin ke...	6.989	7.703	432.5E6	124.3E6	46.769	50.037
22)	Mirex	7.165	8.155	328.5E6	99094528	45.085	49.543

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040219\
Data File : PL046942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2019 02:00
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
4/3/2019 5:56:52 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 05:33:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

