

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040221\
 Data File : PL065823.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Apr 2021 11:54
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 04:55:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 31 15:57:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.277	3.780	134.0E6	136.4E6	51.440	49.686
2)	SA Decachlor...	7.909	8.727	115.7E6	109.5E6	49.968	48.557
Target Compounds							
2)	A alpha-BHC	3.700	4.165	213.3E6	203.1E6	52.162	49.975
3)	MA gamma-BHC...	3.973	4.444	191.0E6	182.4E6	51.589	49.810
4)	MA Heptachlor	4.257	4.930	177.8E6	170.2E6	49.364	48.217
5)	MB Aldrin	4.494	5.225	173.9E6	161.5E6	50.925	48.919
6)	B beta-BHC	4.223	4.628	78561332	86731217	49.289	49.794
7)	B delta-BHC	4.416	4.834	187.3E6	156.7E6	52.066	52.536
8)	B Heptachlo...	4.925	5.612	157.7E6	147.9E6	50.144	47.622
9)	A Endosulfan I	5.252	5.962	145.1E6	132.4E6	50.343	48.241
10)	B gamma-Chl...	5.146	5.842	157.4E6	144.8E6	50.006	48.256
11)	B alpha-Chl...	5.201	5.912	154.6E6	142.3E6	49.700	48.093
12)	B 4,4'-DDE	5.369	6.077	139.7E6	127.9E6	50.644	48.971
13)	MA Dieldrin	5.489	6.217	147.4E6	135.5E6	50.449	48.062
14)	MA Endrin	5.739	6.433	122.3E6	109.5E6	45.281	43.950
15)	B Endosulfa...	6.011	6.650	121.9E6	111.9E6	49.081	47.346
16)	A 4,4'-DDD	5.874	6.567	123.4E6	110.0E6	54.383	52.203
17)	MA 4,4'-DDT	6.107	6.860	93230267	94290608	44.252	42.111
18)	B Endrin al...	6.178	6.774	92601100	90266962	48.940	47.572
19)	B Endosulfa...	6.387	6.995	125.9E6	114.4E6	48.790	47.148
20)	A Methoxychlor	6.652	7.326	54663821	52674644	43.544	41.590
21)	B Endrin ke...	6.865	7.462	137.7E6	126.7E6	50.197	49.295
22)	Mirex	7.037	7.896	99903970	89997202	47.704	46.109

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040221\
Data File : PL065823.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Apr 2021 11:54
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 04:55:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
Quant Title : GC Extractables
QLast Update : Wed Mar 31 15:57:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

