

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040221\
 Data File : PL065858.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Apr 2021 21:15
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 05 05:10:46 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 31 15:57:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.277	3.781	132.5E6	136.1E6	50.867	49.580
2)	SA Decachlor...	7.909	8.726	117.4E6	112.8E6	50.678	50.054
Target Compounds							
2)	A alpha-BHC	3.700	4.166	210.5E6	203.2E6	51.463	50.006
3)	MA gamma-BHC...	3.972	4.445	189.6E6	184.6E6	51.218	50.421
4)	MA Heptachlor	4.257	4.929	159.4E6	162.4E6	44.245	46.003
5)	MB Aldrin	4.493	5.225	172.5E6	166.1E6	50.525	50.309
6)	B beta-BHC	4.223	4.628	78540687	85835664	49.276	49.280
7)	B delta-BHC	4.415	4.834	186.6E6	163.1E6	51.867	54.687
8)	B Heptachlo...	4.925	5.612	158.1E6	152.5E6	50.288	49.118
9)	A Endosulfan I	5.251	5.961	145.9E6	139.2E6	50.623	50.691
10)	B gamma-Chl...	5.146	5.842	157.7E6	152.1E6	50.097	50.710
11)	B alpha-Chl...	5.201	5.912	155.3E6	149.8E6	49.909	50.639
12)	B 4,4'-DDE	5.368	6.076	140.1E6	135.1E6	50.776	51.720
13)	MA Dieldrin	5.488	6.217	147.6E6	143.0E6	50.527	50.743
14)	MA Endrin	5.738	6.432	117.9E6	110.2E6	43.651	44.234
15)	B Endosulfa...	6.011	6.649	123.8E6	118.8E6	49.849	50.285
16)	A 4,4'-DDD	5.874	6.567	126.6E6	118.8E6	55.786	56.352
17)	MA 4,4'-DDT	6.107	6.860	87213298	92117776	41.396	41.141
18)	B Endrin al...	6.177	6.774	95577915	96598160	50.513	50.909
19)	B Endosulfa...	6.387	6.995	128.1E6	121.3E6	49.662	50.007
20)	A Methoxychlor	6.650	7.325	50920670	50846521	40.562	40.146
21)	B Endrin ke...	6.864	7.461	142.0E6	136.4E6	51.790	53.058
22)	Mirex	7.036	7.896	102.8E6	95709195	49.097	49.036

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040221\
Data File : PL065858.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Apr 2021 21:15
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 05 05:10:46 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL033121.M
Quant Title : GC Extractables
QLast Update : Wed Mar 31 15:57:28 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

