

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040224\
 Data File : PL088881.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Apr 2024 13:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 03 01:16:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031324.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 13 15:45:55 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.633	2.851	123.6E6	67760490	46.629	50.843
28)	SA Decachlor...	9.299	8.077	108.3E6	72790016	43.813	53.683
Target Compounds							
2)	A alpha-BHC	4.098	3.365	191.4E6	104.3E6	47.941	51.899
3)	MA gamma-BHC...	4.436	3.701	185.7E6	101.2E6	47.883	52.712
4)	MA Heptachlor	5.036	4.049	179.1E6	96396410	47.165	53.562
5)	MB Aldrin	5.383	4.335	175.8E6	98469161	47.985	54.491
6)	B beta-BHC	4.634	4.003	80716298	43607951	45.514	50.870
7)	B delta-BHC	4.886	4.238	183.4E6	99960622	45.525	52.251
8)	B Heptachlo...	5.815	4.845	170.4E6	88427937	49.603	53.772
9)	A Endosulfan I	6.206	5.221	148.0E6	80117841	48.417	53.297
10)	B gamma-Chl...	6.074	5.099	166.9E6	89909655	50.115	54.283
11)	B alpha-Chl...	6.154	5.164	163.0E6	89934803	49.040	54.180
12)	B 4,4'-DDE	6.329	5.357	144.0E6	81474703	48.187	55.859
13)	MA Dieldrin	6.484	5.490	156.5E6	88923182	48.265	54.751
14)	MA Endrin	6.716	5.769	132.3E6	76097006	48.399	52.793
15)	B Endosulfa...	6.937	6.066	139.4E6	77131148	48.678	53.642
16)	A 4,4'-DDD	6.852	5.919	128.7E6	70603669	50.532	56.693
17)	MA 4,4'-DDT	7.170	6.173	123.4E6	72679554	46.309	54.150
18)	B Endrin al...	7.069	6.248	86682785	60174497	41.518	57.421 #
19)	B Endosulfa...	7.306	6.474	130.6E6	75382047	47.279	54.332
20)	A Methoxychlor	7.650	6.756	68787427	42036959	45.660	52.411
21)	B Endrin ke...	7.797	6.986	143.5E6	90505848	48.548	54.829
22)	Mirex	8.279	7.172	110.8E6	73069989	47.278	54.189
23)	Chlordane-1	0.000	3.874	0	24965	N.D.	0.514 #
24)	Chlordane-2	5.383f	4.472	175.8E6	555094	1337.519	9.143 #
25)	Chlordane-3	6.074	5.099	166.9E6	89909655	335.706	526.802 #
26)	Chlordane-4	6.154	5.164	163.0E6	89934803	271.745	563.714 #
27)	Chlordane-5	0.000	5.867f	0	280986	N.D.	5.540 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040224\
Data File : PL088881.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Apr 2024 13:49
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 03 01:16:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL031324.M
Quant Title : GC Extractables
QLast Update : Wed Mar 13 15:45:55 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.633 min
Delta R.T.: 0.000 min
Response: 123566338
Conc: 46.63 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.851 min
Delta R.T.: -0.001 min
Response: 67760490
Conc: 50.84 ng/ml




#2 alpha-BHC

R.T.: 4.098 min
Delta R.T.: -0.001 min
Response: 191367697
Conc: 47.94 ng/ml



#2 alpha-BHC


R.T.: 3.365 min
Delta R.T.: -0.002 min
Response: 104289468
Conc: 51.90 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.436 min
Delta R.T.: -0.002 min
Response: 185702513
Conc: 47.88 ng/ml

Instrument :
ECD_L
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 3.701 min
Delta R.T.: -0.002 min
Response: 101160334
Conc: 52.71 ng/ml




#4 Heptachlor

R.T.: 5.036 min
Delta R.T.: -0.002 min
Response: 179104982
Conc: 47.16 ng/ml



#4 Heptachlor


R.T.: 4.049 min
Delta R.T.: -0.003 min
Response: 96396410
Conc: 53.56 ng/ml



#5 Aldrin


R.T.: 5.383 min
Delta R.T.: -0.002 min
Response: 175841459
Conc: 47.99 ng/ml

Instrument :
ECD_L
ClientSampleId :




#5 Aldrin

R.T.: 4.335 min
Delta R.T.: -0.003 min
Response: 98469161
Conc: 54.49 ng/ml




#6 beta-BHC

R.T.: 4.634 min
Delta R.T.: -0.001 min
Response: 80716298
Conc: 45.51 ng/ml



#6 beta-BHC


R.T.: 4.003 min
Delta R.T.: -0.003 min
Response: 43607951
Conc: 50.87 ng/ml



#7 delta-BHC


R.T.: 4.886 min
Delta R.T.: -0.001 min
Response: 183385517
Conc: 45.53 ng/ml

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 4.238 min
Delta R.T.: -0.003 min
Response: 99960622
Conc: 52.25 ng/ml




#8 Heptachlor epoxide

R.T.: 5.815 min
Delta R.T.: -0.003 min
Response: 170387105
Conc: 49.60 ng/ml



#8 Heptachlor epoxide


R.T.: 4.845 min
Delta R.T.: -0.004 min
Response: 88427937
Conc: 53.77 ng/ml



#9 Endosulfan I


R.T.: 6.206 min
Delta R.T.: -0.002 min
Response: 147985601
Conc: 48.42 ng/ml

Instrument :
ECD_L
ClientSampleId :




#9 Endosulfan I

R.T.: 5.221 min
Delta R.T.: -0.004 min
Response: 80117841
Conc: 53.30 ng/ml




#10 gamma-Chlordane

R.T.: 6.074 min
Delta R.T.: -0.002 min
Response: 166925214
Conc: 50.12 ng/ml



#10 gamma-Chlordane


R.T.: 5.099 min
Delta R.T.: -0.004 min
Response: 89909655
Conc: 54.28 ng/ml



#11 alpha-Chlordane


R.T.: 6.154 min
Delta R.T.: -0.003 min
Response: 163010975
Conc: 49.04 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 alpha-Chlordane

R.T.: 5.164 min
Delta R.T.: -0.004 min
Response: 89934803
Conc: 54.18 ng/ml




#12 4,4'-DDE

R.T.: 6.329 min
Delta R.T.: -0.003 min
Response: 143967383
Conc: 48.19 ng/ml



#12 4,4'-DDE


R.T.: 5.357 min
Delta R.T.: -0.004 min
Response: 81474703
Conc: 55.86 ng/ml



#13 Dieldrin


R.T.: 6.484 min
Delta R.T.: -0.003 min
Response: 156505866
Conc: 48.26 ng/ml

Instrument :
ECD_L
ClientSampleId :




#13 Dieldrin

R.T.: 5.490 min
Delta R.T.: -0.004 min
Response: 88923182
Conc: 54.75 ng/ml




#14 Endrin

R.T.: 6.716 min
Delta R.T.: -0.003 min
Response: 132308995
Conc: 48.40 ng/ml



#14 Endrin


R.T.: 5.769 min
Delta R.T.: -0.004 min
Response: 76097006
Conc: 52.79 ng/ml



#15 Endosulfan II


R.T.: 6.937 min
Delta R.T.: -0.003 min
Response: 139423535
Conc: 48.68 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 6.066 min
Delta R.T.: -0.004 min
Response: 77131148
Conc: 53.64 ng/ml




#16 4,4'-DDD

R.T.: 6.852 min
Delta R.T.: -0.002 min
Response: 128691727
Conc: 50.53 ng/ml



#16 4,4'-DDD


R.T.: 5.919 min
Delta R.T.: -0.004 min
Response: 70603669
Conc: 56.69 ng/ml



#17 4,4'-DDT


R.T.: 7.170 min
Delta R.T.: -0.003 min
Response: 123434510
Conc: 46.31 ng/ml

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 6.173 min
Delta R.T.: -0.005 min
Response: 72679554
Conc: 54.15 ng/ml




#18 Endrin aldehyde

R.T.: 7.069 min
Delta R.T.: -0.003 min
Response: 86682785
Conc: 41.52 ng/ml



#18 Endrin aldehyde


R.T.: 6.248 min
Delta R.T.: -0.004 min
Response: 60174497
Conc: 57.42 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.306 min
Delta R.T.: -0.003 min
Response: 130550518
Conc: 47.28 ng/ml

Instrument :
ECD_L
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 6.474 min
Delta R.T.: -0.004 min
Response: 75382047
Conc: 54.33 ng/ml




#20 Methoxychlor

R.T.: 7.650 min
Delta R.T.: -0.003 min
Response: 68787427
Conc: 45.66 ng/ml



#20 Methoxychlor


R.T.: 6.756 min
Delta R.T.: -0.004 min
Response: 42036959
Conc: 52.41 ng/ml



#21 Endrin ketone


R.T.: 7.797 min
Delta R.T.: -0.003 min
Response: 143528305
Conc: 48.55 ng/ml

Instrument :
ECD_L
ClientSampleId :




#21 Endrin ketone

R.T.: 6.986 min
Delta R.T.: -0.004 min
Response: 90505848
Conc: 54.83 ng/ml




#22 Mirex

R.T.: 8.279 min
Delta R.T.: -0.002 min
Response: 110843058
Conc: 47.28 ng/ml



#22 Mirex


R.T.: 7.172 min
Delta R.T.: -0.004 min
Response: 73069989
Conc: 54.19 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T.: 4.819 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#23 Chlordane-1

R.T.: 3.874 min
Delta R.T.: 0.000 min
Response: 24965
Conc: 0.51 ng/ml




#24 Chlordane-2

R.T.: 5.383 min
Delta R.T.: 0.028 min
Response: 175841459
Conc: 1337.52 ng/ml



#24 Chlordane-2


R.T.: 4.472 min
Delta R.T.: 0.010 min
Response: 555094
Conc: 9.14 ng/ml



#25 Chlordane-3


R.T.: 6.074 min
Delta R.T.: -0.003 min
Response: 166925214
Conc: 335.71 ng/ml

Instrument :
ECD_L
ClientSampleId :




#25 Chlordane-3

R.T.: 5.099 min
Delta R.T.: -0.003 min
Response: 89909655
Conc: 526.80 ng/ml




#26 Chlordane-4

R.T.: 6.154 min
Delta R.T.: -0.006 min
Response: 163010975
Conc: 271.75 ng/ml



#26 Chlordane-4


R.T.: 5.164 min
Delta R.T.: -0.003 min
Response: 89934803
Conc: 563.71 ng/ml



#27 Chlordane-5


R.T.: 0.000 min
Exp R.T.: 7.019 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#27 Chlordane-5

R.T.: 5.867 min
Delta R.T.: 0.023 min
Response: 280986
Conc: 5.54 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.299 min
Delta R.T.: -0.004 min
Response: 108302417
Conc: 43.81 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.077 min
Delta R.T.: -0.005 min
Response: 72790016
Conc: 53.68 ng/ml