

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040319\
 Data File : PL046967.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2019 17:25
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 01:30:49 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.359	3.978	358.1E6	112.6E6	49.228	49.182
28)	SA Decachlor...	8.037	9.033	426.1E6	118.7E6	50.664	51.200
Target Compounds							
2)	A alpha-BHC	3.788	4.368	602.8E6	162.2E6	51.687	50.478
3)	MA gamma-BHC...	4.065	4.652	543.4E6	151.1E6	50.579	50.475
4)	MA Heptachlor	4.353	5.158	561.6E6	151.8E6	50.934	49.338
5)	MB Aldrin	4.593	5.463	525.0E6	143.4E6	50.900	50.052
6)	B beta-BHC	4.319	4.825	219.3E6	65930499	48.713	49.733
7)	B delta-BHC	4.514	5.041	518.4E6	125.2E6	51.211	53.955
8)	B Heptachlor...	5.030	5.849	495.5E6	130.9E6	50.801	49.619
9)	A Endosulfan I	5.361	6.205	471.0E6	128.9E6	50.090	49.406
10)	B gamma-Chl...	5.253	6.083	513.8E6	140.0E6	50.012	49.345
11)	B alpha-Chl...	5.309	6.155	504.3E6	139.9E6	49.738	49.315
12)	B 4,4'-DDE	5.476	6.314	467.5E6	123.7E6	50.779	49.851
13)	MA Dieldrin	5.601	6.461	507.4E6	133.7E6	50.556	49.311
14)	MA Endrin	5.855	6.679	474.0E6	119.2E6	50.323	49.867
15)	B Endosulfa...	6.129	6.890	448.5E6	121.2E6	50.058	50.239
16)	A 4,4'-DDD	5.988	6.804	383.4E6	100.9E6	50.715	50.140
17)	MA 4,4'-DDT	6.222	7.101	402.1E6	106.6E6	53.083	50.851
18)	B Endrin al...	6.296	7.014	372.5E6	104.5E6	49.697	49.219
19)	B Endosulfa...	6.508	7.237	417.1E6	116.1E6	50.800	49.202
20)	A Methoxychlor	6.767	7.560	215.7E6	64220569	50.750	50.378
21)	B Endrin ke...	6.991	7.707	469.9E6	125.4E6	50.814	50.479
22)	Mirex	7.167	8.158	352.9E6	99951993	48.441	49.972

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040319\
Data File : PL046967.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2019 17:25
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 01:30:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

