

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040319\
 Data File : PL047016.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Apr 2019 05:10
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 4/4/2019 5:20:29 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 05:57:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 22 13:40:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.358	3.976	351.1E6	113.0E6	48.270	49.354
28)	SA Decachlor...	8.033	9.027	404.8E6	115.1E6	48.130	49.644
Target Compounds							
2)	A alpha-BHC	3.787	4.366	572.8E6	164.1E6	49.114	51.067
3)	MA gamma-BHC...	4.064	4.650	522.8E6	152.1E6	48.661	50.810
4)	MA Heptachlor	4.350	5.154	540.5E6	155.6E6	49.021	50.585m
5)	MB Aldrin	4.590	5.459	508.1E6	145.4E6	49.260	50.737
6)	B beta-BHC	4.317	4.823	213.9E6	67911021	47.507	51.227
7)	B delta-BHC	4.512	5.038	505.8E6	130.1E6	49.968	56.080
8)	B Heptachlor...	5.027	5.844	478.9E6	135.8E6	49.096	51.495m
9)	A Endosulfan I	5.359	6.202	456.0E6	128.7E6	48.493	49.320
10)	B gamma-Chl...	5.250	6.080	497.0E6	140.4E6	48.383	49.468
11)	B alpha-Chl...	5.306	6.152	488.8E6	139.5E6	48.212	49.166
12)	B 4,4'-DDE	5.474	6.310	439.4E6	122.3E6	47.720	49.276
13)	MA Dieldrin	5.598	6.457	489.7E6	133.6E6	48.796	49.283m
14)	MA Endrin	5.851	6.676	436.2E6	112.7E6	46.306	47.139
15)	B Endosulfa...	6.126	6.886	430.4E6	120.0E6	48.040	49.737
16)	A 4,4'-DDD	5.984	6.800	374.7E6	101.1E6	49.564	50.246
17)	MA 4,4'-DDT	6.219	7.097	366.7E6	102.4E6	48.420	48.823m
18)	B Endrin al...	6.293	7.010	356.8E6	104.4E6	47.600	49.130
19)	B Endosulfa...	6.504	7.233	394.1E6	114.0E6	48.000	48.338
20)	A Methoxychlor	6.764	7.556	197.4E6	60824925	46.439	47.714
21)	B Endrin ke...	6.987	7.702	449.3E6	124.4E6	48.588	50.054
22)	Mirex	7.163	8.153	328.5E6	98466550	45.088	49.229

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040319\
Data File : PL047016.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Apr 2019 05:10
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
4/4/2019 5:20:29 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 05:57:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL032219.M
Quant Title : GC Extractables
QLast Update : Fri Mar 22 13:40:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

