

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040320\
 Data File : PL057775.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Apr 2020 16:47
 Operator : AJ\MA
 Sample : L2108-05MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP-70-71-COMPMS

Manual Integrations
 APPROVED

mohammad
 4/6/2020 1:31:19 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 07:55:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
 Quant Title : GC Extractables
 QLast Update : Sat Apr 04 06:41:11 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.457	3.956	196.2E6	42388329	19.463	19.532
28)	SA Decachlor...	8.205	8.978	189.3E6	41975622	18.367	18.791
Target Compounds							
2)	A alpha-BHC	3.894	4.343	862.2E6	165.4E6	49.632	50.364
3)	MA gamma-BHC...	4.177	4.626	762.5E6	161.4E6	52.456	49.670
4)	MA Heptachlor	4.478	5.127	827.0E6	172.2E6	51.592	51.308
5)	MB Aldrin	4.725	5.430	761.7E6	166.1E6	51.732	51.297
6)	B beta-BHC	4.430	4.798	337.0E6	75762624	53.357	54.561
7)	B delta-BHC	4.633	5.013	755.7E6	116.8E6	49.792	46.845
8)	B Heptachlo...	5.168	5.813	685.2E6	142.9E6	50.986	52.902
9)	A Endosulfan I	5.507	6.168	630.2E6	152.3E6	51.413	51.075
10)	B gamma-Chl...	5.395	6.047	691.4E6	164.1E6	50.910	50.656
11)	B alpha-Chl...	5.453	6.119	666.7E6	164.7E6	50.070	50.336
12)	B 4,4'-DDE	5.617	6.278	678.0E6	154.4E6	51.752	52.141
13)	MA Dieldrin	5.750	6.425	676.6E6	157.8E6	52.013	51.633
14)	MA Endrin	6.007	6.640	617.2E6	129.8E6	52.873	51.936
15)	B Endosulfa...	6.280	6.851	546.5E6	128.2E6	49.402	49.413
16)	A 4,4'-DDD	6.133	6.767	512.4E6	119.8E6	51.271	52.513
17)	MA 4,4'-DDT	6.373	7.063	539.2E6	116.1E6	53.123	52.139
18)	B Endrin al...	6.449	6.975	468.8E6	111.2E6	50.137	50.063
19)	B Endosulfa...	6.661	7.196	520.3E6	122.2E6	51.210	51.511m
20)	A Methoxychlor	6.918	7.523	280.6E6	64806103	50.906	50.948
21)	B Endrin ke...	7.152	7.665	578.9E6	131.2E6	51.998	53.070
22)	Mirex	7.339	8.115	418.4E6	96612303	51.043	52.546

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL040320\
Data File : PL057775.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Apr 2020 16:47
Operator : AJ\MA
Sample : L2108-05MS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TP-70-71-COMPMS

Manual Integrations
APPROVED

mohammad
4/6/2020 1:31:19 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 07:55:09 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040320.M
Quant Title : GC Extractables
QLast Update : Sat Apr 04 06:41:11 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

